



INSTRUMENT PROCESSING SHEET

Agency Miami Police Department

S/N 80-007516

Florida Department of Law Enforcement

Date In 3/17/2022

DI Completion Date 3/23/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DERR	Date 3/21/2022	Flow Calibration	By	Date															
☒ Annual ☒ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☒ 12V DC Cable Notes: <u>Missing shipping pink bumpers. Contacted CMI DERR</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>204</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17 mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																	
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904 06/08/2023</td> </tr> </tbody> </table>				Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____			
Simulator	Serial #	Lot #/Exp																				
0.050	MP6286	202201C 01/11/2024																				
0.080	MP6287	202201D 01/18/2024																				
0.200	MP6288	202201E 01/18/2024																				
0.080 DGS	N/A	AG115904 06/08/2023																				

Calibration Adjustment				By DERR
Barometric Pressure Gauge 1014 ID # 68639				
Simulator	Serial #	Lot #	Expiration	
0.000	MP5095	N/A	N/A	
0.040	MP5096	21070	03/01/2023	
0.100	MP5098	21080	03/08/2023	
0.200	MP5100	20510	12/03/2022	
0.300	MP5101	21030	02/02/2023	
0.080 DGS	N/A	0052108A2	06/08/2023	
☒ Post Calibration Adjustment Stability Checks <u>DERR 3/23/22</u>				
Simulator	Serial #	Lot #	Expiration	
0.050	MP6286	202201C	01/11/2024	
0.080	MP6287	202201D	01/18/2024	
0.200	MP6288	202201E	01/18/2024	
0.080 DGS	N/A	AG115904	06/08/2023	

Department Inspection		By DERR
Barometric Pressure ID# <u>28199</u>		
Gauge <u>1015</u>	Instrument <u>1014</u>	
Mouth Alcohol Solution Lot # <u>2021-D</u>		
Acetone Stock Solution Lot # <u>2021-C</u>		
Simulator	Serial Number	
0.000	MP6284	
Interferent	MP6285	
0.050	MP6286	
0.080	MP6287	
0.200	MP6288	

Attachments	
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Form 47</u>

Notes/Suggested Service: Instrument was dissabled, by passed AI to bring instrument to ready mode. Instrument was calibrated to bring values closer to nominal. Optical calibration 0.040 solution was high, re-ran solution. Comms failed during upload, I had to go to the level 2 menu to upload. DERR

Tech review identified wrong lot # on cal DGS 00521080A2; Exp 02/05/2023 DERR

☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☒ Return to/Place Into Evidentiary Use
☐ Remain Out of Evidentiary Use
☒ Conduct an Agency Inspection Before Evidentiary Use


Taylor Gutschow 06
10:18:17

Tech Review / Date 03/23/22 Admin Review / Date _____

I have corrected the Department inspection date on the form 47 and checked the boxes in the For Program office use only box as per tech reviewers and Admin reviewer suggestion. DERR 3/29/2022

I added "Compliance not determined AI not Conducted" to the form 40 under the suggestion of the Tech Reviewer but forgot to put a date / initials DERR 3/29/22

I also added a initial to the second page of the optical Calibration DERR 3/29/2022



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007516, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007516</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MIAMI PD</u>	0.050 g/210 L	0.004
Calibration Date:	<u>03/23/2022</u>	0.080 g/210 L	0.004
Calibration Time:	<u>11:03</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/23/2022

Date

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD

Time of Inspection: 11:03

Date of Inspection: 03/23/2022

Serial Number: 80-007516

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.077	0.199	0.080
0.000	0.049	0.077	0.201	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.048	0.078	0.201	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.080	0.201	0.079
0.000	0.049	0.079	0.199	0.079

Standard Deviations	0.0005	0.0009	0.0009	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

03/23/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-007516	Miami Police Department	3/23/2022	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:57 Control Test 0.049 08:57 Air Blank 0.000 08:58 Control Test 0.050 08:59 Air Blank 0.000 08:59 Control Test 0.050 09:00 Air Blank 0.000 09:00 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625	MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:02 Control Test 0.079 09:02 Air Blank 0.000 09:03 Control Test 0.077 09:04 Air Blank 0.000 09:04 Control Test 0.078 09:05 Air Blank 0.000 09:05 Control Test Stats Average 0.0780 Std Dev 0.0010 Rel Std Dev(%) 1.2821	MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:07 Control Test 0.202 09:07 Air Blank 0.000 09:08 Control Test 0.200 09:09 Air Blank 0.000 09:09 Control Test 0.200 09:10 Air Blank 0.000 09:10 Control Test Stats Average 0.2007 Std Dev 0.0012 Rel Std Dev(%) 0.5754	MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:12 Control Test 0.081 09:12 Air Blank 0.000 09:13 Control Test 0.079 09:13 Air Blank 0.000 09:14 Control Test 0.080 09:14 Air Blank 0.000 09:14 Control Test Stats Average 0.0800 Std Dev 0.0010 Rel Std Dev(%) 1.2500
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

MIAMI PD
Intoxilyzer - Alcohol Analyzer
Model 8000
03/23/2022
SN 80-007516
07:47:41
Auto Calibration
Max Power Res Value = 47
Auto Range Res Value = 31

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5100 (-0.0050)
Sample #2 = 1.4560 (0.0340)
Sample #3 = 1.4570 (0.0330)
Sample #4 = 1.4920 (0.0080)
Avg % Abs = 1.4683 (0.0251)
STD DEV = 0.0205 (0.0147)
REL STD DEV = 1.396 (58.924)

Sol Value = 0.040 g/210L xxx
Fit value = 0.1905 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
Sum Io = 12357, Sum Io = 12869

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.6590 (0.0050)
Sample #2 = 0.7260 (0.0080)
Sample #3 = 0.7930 (-0.0310)
Sample #4 = 0.7880 (-0.0310)
Avg % Abs = 0.7690 (-0.0180)
STD DEV = 0.0287 (0.0225)
REL STD DEV = 3.731 (125.093)

Sol Value = 0.000 g/210L xxx
Fit value = 0.0000 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
Sum Io = 12367, Sum Io = 12872

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = -0.0470 (-0.0090)
Sample #2 = 0.0480 (-0.0480)
Sample #3 = 0.0170 (-0.0170)
Sample #4 = 0.0150 (-0.0310)
Avg % Abs = 0.0267 (-0.0220)
STD DEV = 0.0185 (0.0155)
REL STD DEV = 69.383 (48.513)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0630 (-0.0090)
Sample #2 = 0.1420 (-0.0650)
Sample #3 = 0.1270 (-0.0530)
Sample #4 = 0.1140 (-0.0600)
Avg % Abs = 0.1277 (-0.0593)
STD DEV = 0.0140 (0.0060)
REL STD DEV = 10.975 (10.159)

Sol Value = 0.040 g/210L xxx
Fit value = 0.1905 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
Sum Io = 12369, Sum Io = 12881

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8170 (-0.0170)
Sample #2 = 0.7440 (0.0400)
Sample #3 = 0.7310 (0.0690)
Sample #4 = 0.7970 (0.0350)
Avg % Abs = 0.7573 (0.0480)
STD DEV = 0.0350 (0.0184)
REL STD DEV = 4.616 (38.245)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4090 (-0.0110)
Sample #2 = 1.4290 (0.0000)
Sample #3 = 1.4560 (-0.0460)
Sample #4 = 1.4700 (-0.0390)
Avg % Abs = 1.4650 (-0.0283)
STD DEV = 0.0338 (0.0248)
REL STD DEV = 2.306 (87.479)

Sol Value = 0.100 g/210L xxx
Fit value = 0.4762 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
Sum Io = 12357, Sum Io = 12873

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8930 (-0.0160)
Sample #2 = 1.9090 (-0.0160)
Sample #3 = 1.8980 (-0.0210)
Sample #4 = 1.8230 (0.0220)
Avg % Abs = 1.8967 (-0.0050)
STD DEV = 0.0468 (0.0235)
REL STD DEV = 2.494 (470.319)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.4660 (-0.0100)
Sample #2 = 3.4680 (-0.0040)
Sample #3 = 3.4640 (-0.0110)
Sample #4 = 3.4640 (0.0070)
Avg % Abs = 3.4527 (-0.0027)
STD DEV = 0.0232 (0.0091)
REL STD DEV = 0.671 (340.266)

Sol Value = 0.200 g/210L xxx
Fit value = 0.9524 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
Sum Io = 12358, Sum Io = 12875

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6430 (-0.0100)
Sample #2 = 3.6230 (0.0340)
Sample #3 = 3.6210 (0.0360)
Sample #4 = 3.6600 (0.0130)
Avg % Abs = 3.6347 (0.0277)
STD DEV = 0.0220 (0.0127)
REL STD DEV = 0.604 (46.052)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.6060 (-0.0030)
Sample #2 = 6.5920 (0.0450)
Sample #3 = 6.5810 (0.0280)
Sample #4 = 6.6260 (0.0190)
Avg % Abs = 6.5997 (0.0307)
STD DEV = 0.0235 (0.0132)
REL STD DEV = 0.355 (43.055)

Optical Calibration
SN: 80-007516
Agency: Miami PD
Date: 3/23/2022
Quadratic Fit: +/- 0.002g/210L
By: DERR J481

Sol Value = 0.300 g/210L xxx
Fit value = 1.4286 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
Sum Io = 12349, Sum Io = 12868
Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3090 (-0.0050)
Sample #2 = 5.2960 (-0.0120)
Sample #3 = 5.2660 (0.0070)
Sample #4 = 5.3010 (0.0080)
Avg % Abs = 5.2877 (0.0010)
STD DEV = 0.0189 (0.0113)
REL STD DEV = 0.358 (1126.943)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.5530 (0.0050)
Sample #2 = 9.5120 (0.0230)
Sample #3 = 9.5020 (0.0380)
Sample #4 = 9.5450 (0.0360)
Avg % Abs = 9.5197 (0.0233)
STD DEV = 0.0225 (0.0081)
REL STD DEV = 0.236 (25.189)

Auto Cal Data >>>>
Channel 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.027
Std Dev = 0.02 Rel Std Dev = 69.38
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.769
Std Dev = 0.03 Rel Std Dev = 3.73
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.877
Std Dev = 0.05 Rel Std Dev = 2.49
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.635
Std Dev = 0.02 Rel Std Dev = 0.60
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.288
Std Dev = 0.02 Rel Std Dev = 0.36
Zero Order Coef = -50.41
First Order Coef = 2481.24
Second Order Coef = 43.26
Standard Deviation = 15.943202

Channel 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.128
Std Dev = 0.01 Rel Std Dev = 10.98
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.465
Std Dev = 0.03 Rel Std Dev = 2.31
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.453
Std Dev = 0.02 Rel Std Dev = 0.57
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.600
Std Dev = 0.02 Rel Std Dev = 0.46
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.520
Std Dev = 0.02 Rel Std Dev = 0.24
Zero Order Coef = -155.47
First Order Coef = 1367.30
Second Order Coef = 15.61
Standard Deviation = 20.861874

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.0003
0.040 0.040 0.0005
0.100 0.100 0.0001
0.200 0.200 -0.0003
0.300 0.300 0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0004
0.040 0.040 0.0005
0.100 0.100 0.0002
0.200 0.201 -0.0005
0.300 0.300 0.0002

Sol Value = 0.080 g/210L xxx
Fit value = 0.3810 mg/l %xxx
Samples Taken = 4, Discarded = 1

xxxx CHANNEL 1

Sample #1 = 3204.00
Sample #2 = 3232.00
Sample #3 = 3243.00
Sample #4 = 3271.00

Average Result = 3248.6667
STD DEV = 20.1080
REL STD DEV = 0.619

xxxxxxxx

xxxx CHANNEL 2

Sample #1 = 3233.00
Sample #2 = 3252.00
Sample #3 = 3245.00
Sample #4 = 3269.00

Average Result = 3255.3333
STD DEV = 12.3423
REL STD DEV = 0.379

xxxxxxxx

Dry Gas H2O Adjust Results xxxxxxxx
Barometric Pressure = 1014
3 um H2O Adjust (mg/l x 10,000) = 561
9 um H2O Adjust (mg/l x 10,000) = 554
xxxx AUTO CAL PASS

Optical Calibration Continued	
SN:	80-007516
Agency:	Miami PD
Date:	3/23/2022
Quadratic Fit:	+/- 0.002g/210L
By:	DERR <i>[Signature]</i>

0.040 solution values were high, re-ran solution and proceeded with the analysis. DERR

3/29/22

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007516	Miami Police Department	3/21/2022	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 03/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:41 Control Test 0.048 08:42 Air Blank 0.000 08:42 Control Test 0.047 08:43 Air Blank 0.000 08:44 Control Test 0.048 08:44 Air Blank 0.000 08:45 Control Test Status Average 0.0477 Std Dev 0.0006 Rel Std Dev(X) 1.2112	MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 03/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:46 Control Test 0.078 08:47 Air Blank 0.000 08:48 Control Test 0.077 08:48 Air Blank 0.000 08:49 Control Test 0.077 08:49 Air Blank 0.000 08:50 Control Test Status Average 0.0773 Std Dev 0.0006 Rel Std Dev(X) 0.7466	MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 03/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:51 Control Test 0.198 08:52 Air Blank 0.000 08:52 Control Test 0.198 08:53 Air Blank 0.000 08:54 Control Test 0.197 08:54 Air Blank 0.000 08:55 Control Test Status Average 0.1977 Std Dev 0.0006 Rel Std Dev(X) 0.2921	MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 03/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:56 Control Test 0.076 08:57 Air Blank 0.000 08:57 Control Test 0.075 08:57 Air Blank 0.000 08:58 Control Test 0.075 08:58 Air Blank 0.000 08:59 Control Test Status Average 0.0753 Std Dev 0.0006 Rel Std Dev(X) 0.7664
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency:

Time of Inspection: 08:12

Date of Inspection: 03/21/2022

Serial Number: 80-007516

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

Time-Date changed. Time changed. Instrument was in "disabled mode monthly AI expired". Selected AI to bring instrument to ready mode.

Compliance not determined AI not conducted. 3/29/2022 DELL

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E. Reyes-Rivera Signature and Printed Name

DAVID E REYES-RIVERA

03/21/2022
Date



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007516

OWNING AGENCY: Miami Police Department

DATE OF DEPARTMENT INSPECTION: 3/24/2022 ³ ~~DEAL~~ 3/23/22

AGENCY INSPECTOR: Officer Jose Pastor

ADDRESS: 2200 Flagler St

CITY, STATE, ZIP: Miami, FL 33135

TELEPHONE NUMBER: (305) 603-6537

FAX NUMBER: _____

EMAIL ADDRESS (if available): 5533@miami-police.org

For Program Office Use Only:

Date:
03/24/22

Registration Issued

☒

Instrument Added to Evidentiary Instrument Database

☒

Instrument Added to Monthly Statistics Database

☒

Contact Information Added to Instrument Database

3/29/2022 DERR

3/29/2022 DERR