



Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: St. Lucie County Sheriff's Office Instrument Serial Number: 788, 792, and 794

Agency Notification	Records Audit:
Agency Inspector/Contact: <u>David Caglioni</u>	The following records were audited: <input type="checkbox"/> Instrument Registration <input type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input type="checkbox"/> Other: _____
Date of Notification: <u>6/30/2022</u>	
Comments: <u>Unable to audit records (not stored at the visited location).</u> _____ _____ _____	

Instrument/Area:	Equipment:	Supplies:
<input checked="" type="checkbox"/> Clean/Dry <input checked="" type="checkbox"/> Secure <input checked="" type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input checked="" type="checkbox"/> Proper Number of Simulators <input checked="" type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input checked="" type="checkbox"/> Class A Glassware	<input checked="" type="checkbox"/> Distilled/Deionized Water <input checked="" type="checkbox"/> Mouth Alcohol Solution <input checked="" type="checkbox"/> Acetone Stock Solution <input checked="" type="checkbox"/> Alcohol Reference Solution <input checked="" type="checkbox"/> Dry Gas Standard <input checked="" type="checkbox"/> Mouth Pieces
Comments: <u>Observed the Agency Inspection process. Took one instrument back with me for the annual inspection.</u> <u>Discussed best upload practices. Compliant with Chapter 11D-8, Florida Administrative Code.</u> _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		

Taylor Gutschow

Signature of Alcohol Testing Program Staff Member

Digitally signed by Taylor Gutschow
Date: 2022.07.01 15:28:53 -04'00'

7/1/2022

Date