



Florida Department of  
Law Enforcement

## DEPARTMENT INSPECTOR FIELD NOTES

Agency: Citrus County SO

Instrument Serial Number: 80-000819

Agency Notification	Records Audit:
<b>Agency Inspector/Contact:</b> <u>Mary Bowermaster</u>	<b>The following records were audited:</b> <input checked="" type="checkbox"/> Instrument Registration <input checked="" type="checkbox"/> Agency Inspection Reports <input checked="" type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input checked="" type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis  <input type="checkbox"/> Other: _____
<b>Date of Notification:</b> <u>02/17/2022</u>	
<b>Comments:</b> _____ _____ _____	

Instrument/Area:	Equipment:	Supplies:
<input checked="" type="checkbox"/> Clean/Dry <input checked="" type="checkbox"/> Secure <input checked="" type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input checked="" type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input checked="" type="checkbox"/> Class A Glassware	<input checked="" type="checkbox"/> Distilled/Deionized Water <input checked="" type="checkbox"/> Mouth Alcohol Solution <input checked="" type="checkbox"/> Acetone Stock Solution <input checked="" type="checkbox"/> Alcohol Reference Solution <input checked="" type="checkbox"/> Dry Gas Standard <input checked="" type="checkbox"/> Mouth Pieces
<b>Comments:</b> Access to Intoxilyzer has limited access and requires fingerprint scan to access keys to breath test room. AI and BTOs only have access to keys. All records are kept on site then sent to SAO if requested. Camera with external hard drive in breath testing room. _____ _____ _____ _____ _____ _____ _____ _____ _____		

Richard A Williams

Digitally signed by Richard A Williams  
Date: 2022.02.24 09:36:30 -05'00'

Signature of Alcohol Testing Program Staff Member

Date