

## DEPARTMENT INSPECTOR FIELD NOTES

Agency: Sarasota County She	eriff's Office	Instrument Serial Number: 80-001346	
Agency Notification	Record	s Audit:	
Agency Inspector/Contact: Mike Feltovic  Date of Notification: 7/21/2022	☐ Insti	The following records were audited: Instrument Registration Agency Inspection Reports Breath Test Instrument Repair/Maintenance Records Dry Gas Standard Certificate(s) of Analysis Other: Prior field visit paperwork	
Comments: The breath test operational	procedures and last Agend	ry Inspection date were posted in the area.	
Instrument/Area:	Equipment:	Supplies:	
■ Clean/Dry ■ Secure ■ Limited Access □ Other:	Proper Number of Sim Checked Simulators for Resistant Seal and Pro Temperature Class A Glassware	or Air Leak Mouth Alcohol Solution	
Comments:  Discussed best upload pract	ices. Agency is compliant w	vith Chapter 11D-8, Florida Administrative Code.	
Taylor Guischow	igitally signed by Taylor Gutschow ate: 2022.07.22 14:49:38 -04'00'	7/22/2022 Date	