



Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Hendry County Sheriff's Office

Instrument Serial Numbers: 951, 1207, 6234, and 6386

Agency Notification	Records Audit:
Agency Inspector/Contact: <u>James Hester</u>	The following records were audited: <input checked="" type="checkbox"/> Instrument Registration <input checked="" type="checkbox"/> Agency Inspection Reports <input checked="" type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input checked="" type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input type="checkbox"/> Other: _____
Date of Notification: <u>4/27/2022</u>	
Comments: _____ _____ _____	

Instrument/Area:	Equipment:	Supplies:
<input checked="" type="checkbox"/> Clean/Dry <input checked="" type="checkbox"/> Secure <input checked="" type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input checked="" type="checkbox"/> Proper Number of Simulators <input checked="" type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input checked="" type="checkbox"/> Class A Glassware	<input checked="" type="checkbox"/> Distilled/Deionized Water <input checked="" type="checkbox"/> Mouth Alcohol Solution <input checked="" type="checkbox"/> Acetone Stock Solution <input checked="" type="checkbox"/> Alcohol Reference Solution <input checked="" type="checkbox"/> Dry Gas Standard <input checked="" type="checkbox"/> Mouth Pieces
Comments: Observed AI Hester perform an Agency Inspection in accordance with Chapter 11D-8, FAC. Discussed best practices for uploading and how to become a Breath Test Instructor. Added AI Hester to the list of individuals interested in future BTI courses taught by FDLE. _____ _____ _____ _____ _____ _____ _____ _____ _____		

Taylor Gutschow

Signature of Alcohol Testing Program Staff Member

Digitally signed by Taylor Gutschow
Date: 2022.04.27 13:06:50 -04'00'

4/27/2022

Date