



Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Kennedy Space Center Instrument Serial Number: 80-006621 and 80-006759

Agency Notification	Records Audit:
Agency Inspector/Contact: <u>Karen Joyner</u>	The following records were audited: <input checked="" type="checkbox"/> Instrument Registration <input checked="" type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input checked="" type="checkbox"/> Other: <u>Permitted BTO/AI list</u>
Date of Notification: <u>8/3/2022</u>	
Comments: <u>I met with AI Joyner on 8/3/2022.</u> 	

Instrument/Area:	Equipment:	Supplies:
<input checked="" type="checkbox"/> Clean/Dry <input checked="" type="checkbox"/> Secure <input checked="" type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input checked="" type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input type="checkbox"/> Class A Glassware	<input checked="" type="checkbox"/> Distilled/Deionized Water <input checked="" type="checkbox"/> Mouth Alcohol Solution <input checked="" type="checkbox"/> Acetone Stock Solution <input checked="" type="checkbox"/> Alcohol Reference Solution <input checked="" type="checkbox"/> Dry Gas Standard <input checked="" type="checkbox"/> Mouth Pieces
Comments: 80-006759 was taken out of service by the Agency Inspector. It will be taken back to ATP for evaluation/calibration. The AI indicated numerous Interferent Detect messages during the last Agency Inspection. Worked with Agency Inspector to connect and upload data to FDLE. I changed the settings in the Intoxilyzer 8000 (80-006621) so the AI would no longer need to type the number when performing a Comms Transfer. 		

Date: 2022.08.08
11:59:14 -04'00'

Signature of Alcohol Testing Program Staff Member

Date