

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Florida Highway Patrol, Troop C Instrument Serial Number: 80-006632			
Agency Notification	Records Audit:		
The following records we			
Agency Inspector/Contact: William Smith and Anthony Dobosiewicz		☐ Instrument Registration	
	☐ Agency Inspection R	-	
Date of Notification:		☐ Breath Test Instrument Repair/Maintenance Records ☐ Dry Gas Standard Certificate(s) of Analysis	
2/24/2022			
	□ Other:		
Comments:			
Luckinin and Austri	F	Compliant	
Instrument/Area: Clean/Dry	Equipment: Proper Number of Simulators	Supplies: Distilled/Deionized Water	
Secure	☐ Checked Simulators for Air Leak	Mouth Alcohol Solution	
Limited Access	Resistant Seal and Proper	Acetone Stock Solution	
☐ Other:	Temperature Class A Glassware	■ Alcohol Reference Solution	
	Class A Glassware	■ Dry Gas Standard■ Mouth Pieces	
Comments:			
Instrument 6632 was sent to FDLE in December 2021 for an Ambient Fail and again in January 2022			
for a Control Outside Tolerance. Neither exception message could be duplicated at FDLE. Al reported			
6632 failed its February Agency Inspection due to an inability to detect acetone during the Interferent			
Test. I went on-site and observed the AI set up instrument/simulators. I noticed the AI did not check			
for a leak-resistant seal, so I reminded him it was part of the Form 39- Agency Inspection Procedure			
and demonstrated how to test for a good seal. Ran a Stability Test on the Alcohol Free Solution and			
noticed a lack of bubbles inside the simulator jar during sample collection. The same thing occurred			
when a Stability Test was run on the Interferent Solution. 6632 is able to pull samples from the DGS			
cylinder without issue. May be related to rough shipping, as this did not occur at FDLE. 6632 will			
be sent to repair. Al recalled a lack of bubbles in the simulator jar during his recent failed inspection,			
so I reprinted the Form 40 (he did not print before upload) and instructed him to add a remark.			
	ned by Taylor Gutschow 02.25 11:55:01 -05'00'	2/25/2022	
Signature of Alcohol Testing Pro	ogram Staff Member	Date	