



Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Kennedy Space Center Instrument Serial Number: 80-006621 and 80-006759

Agency Notification	Records Audit:
Agency Inspector/Contact: <u>Karen Joyner</u>	The following records were audited: ☒ Instrument Registration ☒ Agency Inspection Reports ☐ Breath Test Instrument Repair/Maintenance Records ☐ Dry Gas Standard Certificate(s) of Analysis ☒ Other: <u>Permitted BTO/AI list</u>
Date of Notification: <u>8/3/2022</u>	
Comments: <u>I met with AI Joyner on 8/3/2022.</u> 	

Instrument/Area:	Equipment:	Supplies:
☒ Clean/Dry ☒ Secure ☒ Limited Access ☐ Other: _____	☒ Proper Number of Simulators ☐ Checked Simulators for Air Leak Resistant Seal and Proper Temperature ☐ Class A Glassware	☒ Distilled/Deionized Water ☒ Mouth Alcohol Solution ☒ Acetone Stock Solution ☒ Alcohol Reference Solution ☒ Dry Gas Standard ☒ Mouth Pieces
Comments: <u>80-006759 was taken out of service by the Agency Inspector. It will be taken back to ATP for evaluation/calibration. The AI indicated numerous Interferent Detect messages during the last Agency Inspection.</u> <u>Worked with Agency Inspector to connect and upload data to FDLE. I changed the settings in the Intoxilyzer 8000 (80-006621) so the AI would no longer need to type the number when performing a Comms Transfer.</u> 		

Date: 2022.08.08
11:59:14 -04'00'

Signature of Alcohol Testing Program Staff Member

Date