



Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Desoto County Sheriff's Office Instrument Serial Number: 7508 and 7509

Agency Notification	Records Audit:
Agency Inspector/Contact: <u>Joshua Pitts</u>	The following records were audited: ☒ Instrument Registration ☒ Agency Inspection Reports ☐ Breath Test Instrument Repair/Maintenance Records ☒ Dry Gas Standard Certificate(s) of Analysis ☐ Other: _____
Date of Notification: <u>6/7/2022</u>	
Comments: <u>No repair records on 7508 or 7509 (new instruments).</u> _____ _____ _____	

Instrument/Area:	Equipment:	Supplies:
☒ Clean/Dry ☒ Secure ☒ Limited Access ☐ Other: _____	☒ Proper Number of Simulators ☐ Checked Simulators for Air Leak Resistant Seal and Proper Temperature ☒ Class A Glassware	☒ Distilled/Deionized Water ☒ Mouth Alcohol Solution ☒ Acetone Stock Solution ☒ Alcohol Reference Solution ☒ Dry Gas Standard ☒ Mouth Pieces
Comments: <u>Discussed best upload practices. Picked up instrument for annual inspection. Compliant with Chapter</u> <u>11D-8, Florida Administrative Code.</u> _____ _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		

Taylor Gutschow

Signature of Alcohol Testing Program Staff Member

Digitally signed by Taylor Gutschow
Date: 2022.06.13 10:36:49 -04'00'

6/13/2022

Date