

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO

Time of Inspection: 12:11

Date of Inspection: 01/17/2023

Serial Number: 80-000744

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1251438 Exp: 02/25/2023
0.000	0.048	0.078	0.199	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.049	0.080	0.201	0.081

Number of Simulators Used: 5


Remarks:

IN COMPLIANCE

POWER WENT OUT DURING INSPECTION, REDID INSPECTION.

The above instrument complies (☒) does not comply (☐) with Chapter 11B-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11B-8, FAC.

 176 BLSO
 Signature and Printed Name SAMUEL L. SHELLEY
 Date 01/17/2023