

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 23:09

Date of Inspection: 11/28/2023

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08223080A3 Exp: 06/05/2025
0.000	0.047	0.078	0.195	0.083
0.000	0.048	0.078	0.197	0.082
0.000	0.048	0.078	0.196	0.083

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SA PS 1412

Signature and Printed Name

PATRICK M BAKER

11/28/2023

Date

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
11/28/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	21:45
Control Test	0.000	21:46
Air Blank	0.000	21:46
Control Test	0.000	21:47
Air Blank	0.000	21:47
Control Test	0.000	21:48
Air Blank	0.000	21:48
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
11/28/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	21:50
Control Test	INT*	21:50
Air Blank	0.000	21:51
Control Test	INT*	21:52
Air Blank	0.000	21:52
Control Test	INT*	21:53
Air Blank	0.000	21:53
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*Interferent Detect

SA B1463

Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
11/28/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	21:54
Control Test	0.047	21:55
Air Blank	0.000	21:55
Control Test	0.047	21:56
Air Blank	0.000	21:57
Control Test	0.047	21:57
Air Blank	0.000	21:58
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Sgt B. 11/28/23
Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
11/28/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	22:03
Control Test	0.077	22:04
Air Blank	0.000	22:04
Control Test	0.077	22:05
Air Blank	0.000	22:06
Control Test	0.078	22:06
Air Blank	0.000	22:07
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	


Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
11/28/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	22:08
Control Test	0.194	22:09
Air Blank	0.000	22:10
Control Test	0.194	22:10
Air Blank	0.000	22:11
Control Test	0.194	22:11
Air Blank	0.000	22:12
Control Test Stats		
Average	0.1940	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Sgt P51463
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Serial Number: 80-000812

Time of Inspection: 20:35

Date of Inspection: 10/30/2023

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 08223080A3 Exp: 06/05/2025
0.000	0.048	0.076	0.194	0.082
0.000	0.048	0.078	0.196	0.082
0.000	0.049	0.078	0.195	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

10/30/2023

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 19:13

Date of Inspection: 08/13/2023

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08223080A3 Exp: 06/05/2025
0.000	0.048	0.077	0.195	0.081
0.000	0.048	0.078	0.195	0.081
0.000	0.048	0.078	0.196	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

PATRICK M BAKER

08/13/2023
 Date

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
08/13/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:10
Control Test	0.000	18:10
Air Blank	0.000	18:11
Control Test	0.000	18:11
Air Blank	0.000	18:12
Control Test	0.000	18:13
Air Blank	0.000	18:13
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

 1405
Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
08/13/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:15
Control Test	INT*	18:15
Air Blank	0.000	18:16
Control Test	INT*	18:16
Air Blank	0.000	18:17
Control Test	INT*	18:18
Air Blank	0.000	18:18
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*Interferent Detect


Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
08/13/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:19
Control Test	0.047	18:20
Air Blank	0.000	18:20
Control Test	0.047	18:21
Air Blank	0.000	18:22
Control Test	0.048	18:22
Air Blank	0.000	18:23
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	


Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
08/13/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:24
Control Test	0.077	18:24
Air Blank	0.000	18:25
Control Test	0.078	18:26
Air Blank	0.000	18:26
Control Test	0.077	18:27
Air Blank	0.000	18:27
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	


Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
08/13/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:28
Control Test	0.194	18:29
Air Blank	0.000	18:30
Control Test	0.195	18:30
Air Blank	0.000	18:31
Control Test	0.194	18:32
Air Blank	0.000	18:32
Control Test Stats		
Average	0.1943	
Std Dev	0.0006	
Rel Std Dev(%)	0.2971	



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 19:01

Date of Inspection: 07/19/2023

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30822080A1 Exp: 12/05/2024
0.000	0.047	0.077	0.195	0.083
0.000	0.047	0.077	0.195	0.082
0.000	0.047	0.077	0.195	0.083

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 PATRICK M BAKER
 

Signature and Printed Name

07/19/2023

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 19:41

Date of Inspection: 07/19/2023

Serial Number: 80-000814

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30822080A1 Exp: 12/05/2024
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.079	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CP Patrick M Baker 1463

Signature and Printed Name

PATRICK M BAKER

BRAD E COLLINS

2480193

07/19/2023

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 18:45

Date of Inspection: 07/19/2023

Serial Number: 80-005249

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08223080A3 Exp: 06/05/2025
0.000	0.048	0.078	0.198	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.050	0.079	0.200	0.077

Number of Simulators Used: 5

Remarks:

PASSED INSPECTION

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick M Baker

1968 Sgt B Baker

PATRICK M BAKER

BRAND R COLLITO

Signature and Printed Name

07/19/2023

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 18:20

Date of Inspection: 05/11/2023

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30822080A1 Exp: 12/05/2024
0.000	0.047	0.076	0.195	0.081
0.000	0.047	0.076	0.195	0.082
0.000	0.048	0.076	0.195	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BRAD R COLLITO

Signature and Printed Name

05/11/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 08:51

Date of Inspection: 04/11/2023

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30822080A1 Exp: 12/05/2024
0.000	0.048	0.078	0.197	0.082
0.000	0.048	0.077	0.197	0.083
0.000	0.048	0.078	0.197	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature

BRAD R COLLITO

Signature and Printed Name

04/11/2023

Date

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 13:50

Date of Inspection: 03/17/2023

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30822080A1 Exp: 12/05/2024
0.000	0.047	0.076	0.195	0.082
0.000	0.046	0.076	0.196	0.083
0.000	0.047	0.077	0.196	0.083

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Collito

BRAD R COLLITO

Signature and Printed Name

03/17/2023

Date

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 14:59

Date of Inspection: 01/14/2023

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:15922080A5 Exp: 08/05/2024
0.000	0.048	0.078	0.198	0.082
0.000	0.048	0.079	0.199	0.083
0.000	0.048	0.079	0.200	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BR 1193

BRAD R COLLITO

Signature and Printed Name

01/14/2023

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 10:03

Date of Inspection: 02/11/2023

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:15922080A5 Exp: 08/05/2024
0.000	0.049	0.078	0.198	0.083
0.000	0.049	0.079	0.198	0.082
0.000	0.049	0.079	0.198	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BRAD R. COLLITO

BRAD R COLLITO

Signature and Printed Name

02/11/2023
Date

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO
Time of Inspection: 14:16

Date of Inspection: 01/14/2023

Serial Number: 80-005249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:15922080A5 Exp: 08/05/2024
0.000	0.048	0.079	0.197	0.082
0.000	0.048	0.080	0.199	0.082
0.000	0.049	0.081	0.199	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BC 1193

BRAD R COLLITO

Signature and Printed Name

01/14/2023
Date