



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Sarasota CSO

Instrument Serial Number: 80-000861

AGENCY INSPECTION DISCREPANCY

Agency Inspector: Timothy Brenckle

Inspection Date: 02/24/2023

☒ **A test was repeated without reason for repeat or the corrective action given**

- ☐ Alcohol Free/Mouth Alcohol ☐ Alcohol Free Test ☐ Interferent Detect
☐ 0.050 g/210L ☐ 0.080 g/210L ☒ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ **Agency Inspection not conducted for calendar month**

☐ **Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection**

- ☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ **Lot Number and/or Expiration Date for Standards entered incorrectly**

- ☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☒ **Other Electronic Data Review/Comments:** Include a remark describing the corrective action that was taken prior to repeating the 0.20 Test.

REQUESTED CORRECTIVE ACTION

☒ **Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector**

☐ **Provide written explanation regarding referenced item(s) to Department Inspector**

☐ **Remove instrument from evidentiary use until otherwise directed by the Department**

☐ **Other:**

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2023.05.17 08:55:46
-04'00'

Signature of Alcohol Testing Program Staff Member

05/17/2023

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 20:00

Date of Inspection: 02/24/2023

Serial Number: 80-000861
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:15922080A4 Exp: 08/05/2024
0.000	0.048	0.078	0.200 / 0.199	0.078
0.000	0.048	0.078	0.200 / 0.200	0.078
0.000	0.048	0.078	/ 0.200	0.078

Number of Simulators Used: 5

Remarks:

20: RFI Detect.

Deputy came in with radio on. Removed deputy from room before resuming test.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

0/5  2922

TIMOTHY M BRECKLE

Signature and Printed Name

02/24/2023
Date