

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 09:08

Date of Inspection: 03/07/2023

Serial Number: 80-000870
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046 / 0.045			
0.000	0.045 / 0.044			
0.000	0.044 / 0.043			

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: 0.05 CONTROL OUTSIDE OF TOL.

Cleaning lady was mopping the KV Station with an alcohol based cleaner. Vented area and waited for mopping solution to dry. Performed second Agency Inspection without any issues.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jason T Keith 7/14/2013

Signature and Printed Name

JASON T KEITH

03/07/2023
Date