



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Miami PD

Instrument Serial Number: 80-000873

AGENCY INSPECTION DISCREPENCY

Agency Inspector: Angel Fernandez

Inspection Date: 06/19/2023

☐ A test was repeated without reason for repeat or the corrective action given

☐ Alcohol Free/Mouth Alcohol ☐ Alcohol Free Test ☐ Interferent Detect

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ Agency Inspection not conducted for calendar month

☐ Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☒ Lot Number and/or Expiration Date for Standards entered incorrectly

☐ 0.050 g/210L ☐ 0.080 g/210L ☒ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ Other Electronic Data Review/Comments:

REQUESTED CORRECTIVE ACTION

☒ Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector

☐ Provide written explanation regarding referenced item(s) to Department Inspector

☐ Remove instrument from evidentiary use until otherwise directed by the Department

☐ Other:

Taylor Gutschow

Digitally signed by Taylor
Gutschow

Date: 2023.07.06 12:24:11 -04'00'

07/06/2023

Signature of Alcohol Testing Program Staff Member

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD

Time of Inspection: 13:11

Date of Inspection: 06/19/2023

Serial Number: 80-000873

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) <i>AF</i> Lot#: 202106B <i>AF</i> Exp: 06/10/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1506449 Exp: 06/10/2024
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.080

Number of Simulators Used: 5

Remarks:

Amended 0.20g/210L test lot# 202106B exp 6/10/2024

28394
9/8/23

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

28394

ANGEL FERNANDEZ

Signature and Printed Name

06/19/2023
Date