

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Miami Beach Police Department

Serial Number: 80-000879

Time of Inspection: N/A

Date of Inspection: N/A

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted	N	A
Diagnostic Check (Pre-Inspection): OK	N	A
Alcohol Free Subject Test: 0.000	N	A
Mouth Alcohol Test: Slope Not Met	N	A
Interferent Detect Test: Interferent Detect	N	A
Diagnostic Check (Post-Inspection): OK	N	A

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A

Number of Simulators Used: N/A

Remarks: I am writing to inform you that Intoxilyzer 80-000879 was taken out of service on Thursday, February 23rd, 2023. The Intoxilyzer is being sent for service to CMI due to an issue with it not wanting to power on. As a result, we are unable to complete an agency inspection or run any subject tests. Due to this power issue, an agency inspection was not conducted for February 2023 and all the data was not uploaded to FDLE.

Taylor Gutschow (FDLE) and the Miami SAO DUI Chiefs have been advised. I will advise once the Intoxilyzer is back in service. Please advise if there are any questions regarding this matter.

The above instrument complies () does not comply (☒) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PA Socarras
#501

Pedro A. Socarras

Signature and Printed Name

02/23/2023
Date