

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD  
Time of Inspection: 03:38

Date of Inspection: 03/02/2023

Serial Number: 80-000884  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: \_\_\_\_\_

Remarks:

1 accidentally hit "inspection test" after  
printing passed ATP form 40 (insp @ 0335  
on 3/2/2023).

1st insp passed. 2nd entry of inspection test  
waived. - *[Signature]* 3195/A1

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]* 3195

RICHARD L CLOSIUS

Signature and Printed Name

03/02/2023  
Date

R. MYRIL

*[Signature]* #5383