

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GLADES COUNTY S.O.
Time of Inspection: 11:02

Date of Inspection: 08/21/2023

Serial Number: 80-000947
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06022080A4 Exp: 04/05/2024
0.000	0.047	0.078	0.201	0.080
0.000	0.047	0.079	0.201	0.079
0.000	0.047	0.079	0.202	0.078

Number of Simulators Used: 4

Remarks:

A F / M A: USER ERROR Improper Sample.

*user got distracted forgot what step
I was on*

471

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

TRAVIS J ARNOLD

08/21/2023
Date