

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 13:36

Date of Inspection: 07/14/2021

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1094981 Exp: 02/22/2022
0.000	0.047	0.082	0.194	0.069 / 0.076
0.000	0.048	0.081	0.194	0.069 / 0.076
0.000	0.047	0.081	0.194	0.069 / 0.076

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance LOW TANK PRESSURE

TANK AT 1/2 - CHANGED TANK WITH
FULL TANK - NOW WITHIN TOLERANCE.

(Signature)

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

(Signature) #17979

KYLE DUFFEY

Signature and Printed Name

07/14/2021
Date