

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Jacksonville Sheriff's Office

Serial Number: 80-001069

Time of Inspection: 1000

Date of Inspection: 12/15/2023

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		NO
Diagnostic Check (Pre-Inspection): OK		NO
Alcohol Free Subject Test: 0.000		NO
Mouth Alcohol Test: Slope Not Met		NO
Interferent Detect Test: Interferent Detect		NO
Diagnostic Check (Post-Inspection): OK		NO

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: 0

Instrument would not come to (Ready Mode). Instrument removed from Evidentiary use at the PTFD. Unable to perform Agency Inspection due to Diagnostic Failure of DSP system in Instrument #80-001069. Notified my Departmental Inspector of Issue and removal of Instrument from Evidentiary use. Sent back to manufacturer(CMI)for inspection and maintenance. Then FDLE for Departmental Inspection.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
69016

Robert D Lightsey

Signature and Printed Name

12/15/2023  
Date