

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BOYNTON BEACH PD
Time of Inspection: 11:05

Date of Inspection: 11/29/2023

Serial Number: 80-001190
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:220213 Exp: 02/10/2025
0.000	0.049	0.079	0.164 / 0.200	0.081
0.000	0.049	0.079	0.200 / 0.199	0.081
0.000	0.049	0.078	0.199 / 0.196	0.081

Number of Simulators Used: 5

Remarks:

A F / M A:GAVE READING 0.062, NO SNM. 20: Control Outside LIDNOT ON TIGHT

- * THIS IS FINAL INSPECTION REPORT - SPOKE WITH FDLE TAYLOR GUTSCHOW IN REFERENCE TO PROCEDURES*
- ORIGINALLY SWISHED MOUTH ALCOHOL PRIOR TO FIRST BLOW GIVING A READING INSTEAD OF 0.000 - STARTED NEW TEST*
 - 1ST MOUTH ALCOHOL BLEW READING, 2ND TIME FOR 0.000 BREATH EXPIRED*
 - CALLED FDLE GUTSCHOW ADVISED TO WAIT 30 MINS AND RESTART*
 - INTERFERENCE TEST #1 ALL 0.000, SWITCHED ACETONE SOLUTION. TEST #2 2 OF 3 INT*
 - CALLED FDLE GUTSCHOW ADVISED TO WAIT 15 MINS AND RESTART*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brittany L Drury
Signature and Printed Name

BRITTANY L DRURY

11/29/2023
Date