



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: West Palm Beach PD

Instrument Serial Number: 80-001235

AGENCY INSPECTION DISCREPENCY

Agency Inspector: James Ingrassia

Inspection Date: 06/23/2023

☒ A test was repeated without reason for repeat or the corrective action given

☐ Alcohol Free/Mouth Alcohol ☒ Alcohol Free Test ☐ Interferent Detect

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ Agency Inspection not conducted for calendar month

☐ Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ Lot Number and/or Expiration Date for Standards entered incorrectly

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ Other Electronic Data Review/Comments:

REQUESTED CORRECTIVE ACTION

☒ Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector

☐ Provide written explanation regarding referenced item(s) to Department Inspector

☐ Remove instrument from evidentiary use until otherwise directed by the Department

☐ Other:

Taylor Gutschow

Digitally signed by Taylor

Gutschow

Date: 2023.09.12 13:13:39 -04'00'

09/12/2023

Signature of Alcohol Testing Program Staff Member

Date

Amended 9/13/23

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT & INTOXILYZER 8000

Agency: WEST PALM BEACH PD
Time of Inspection: 12:09

Date of Inspection: 06/23/2023

Serial Number: 80-001235
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				
0.009				
0.000				

Number of Simulators Used: 4 _____

Remarks:

00: Control Outside Tolerance. CONTROL OUTSIDE TOLERANCE ON 0.

inspection aborted - wet control test
on alcohol free done & complied.
inspection restarted *[Signature]* 9/13/23

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAMES A INGRASSIA

Signature and Printed Name

06/23/2023
Date



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: West Palm Beach PD

Instrument Serial Number: 80-001235

AGENCY INSPECTION DISCREPENCY

Agency Inspector: James Ingrassia

Inspection Date: 05/24/2023

☒ **A test was repeated without reason for repeat or the corrective action given**

☐ Alcohol Free/Mouth Alcohol ☒ Alcohol Free Test ☐ Interferent Detect

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ **Agency Inspection not conducted for calendar month**

☐ **Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection**

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ **Lot Number and/or Expiration Date for Standards entered incorrectly**

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☒ **Other Electronic Data Review/Comments:** The Alcohol-Free Test was repeated. Describe the corrective action taken prior to repeating the test.

REQUESTED CORRECTIVE ACTION

☒ **Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector**

☐ **Provide written explanation regarding referenced item(s) to Department Inspector**

☐ **Remove instrument from evidentiary use until otherwise directed by the Department**

☐ **Other:**

Taylor Gutschow

Digitally signed by Taylor

Gutschow

Date: 2023.06.15 13:49:04 -04'00'

06/15/2023

Signature of Alcohol Testing Program Staff Member

Date

AMENDED* 6/15/23

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT @ INTOXILYZER 8000

Agency: WEST PALM BEACH PD
Time of Inspection: 14:30

Date of Inspection: 05/24/2023

Serial Number: 80-001235
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 02422080A1 Exp: 02/05/2024
/ 0.000	0.048	0.078	0.196	0.079
/ 0.000	0.048	0.078	0.197	0.079
/ 0.000	0.048	0.079	0.198	0.079

Number of Simulators Used: 4

Remarks:

00: RFI Detect.

SOURCE OF RFI REMOVED FROM AREA; TEST REPEATS (JH) 6/15/23

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 JAMES A INGRASSIA
Signature and Printed Name

05/24/2023
Date