



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Indian Harbour PD

Instrument Serial Number: 80-001263

AGENCY INSPECTION DISCREPENCY

Agency Inspector: Matthew Wimmer

Inspection Date: 08/29/2023

☐ A test was repeated without reason for repeat or the corrective action given

☐ Alcohol Free/Mouth Alcohol ☐ Alcohol Free Test ☐ Interferent Detect

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ Agency Inspection not conducted for calendar month

☐ Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☒ Lot Number and/or Expiration Date for Standards entered incorrectly

☐ 0.050 g/210L ☐ 0.080 g/210L ☒ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☒ Other Electronic Data Review/Comments: The 0.20 ARS lot number is mistyped (missing a digit).

REQUESTED CORRECTIVE ACTION

☒ Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector

☐ Provide written explanation regarding referenced item(s) to Department Inspector

☐ Remove instrument from evidentiary use until otherwise directed by the Department

☐ Other:

Taylor Gutschow

Digitally signed by Taylor

Gutschow

Date: 2023.10.31 12:26:50 -04'00'

10/31/2023

Signature of Alcohol Testing Program Staff Member

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN HARBOUR PD
Time of Inspection: 09:14

Date of Inspection: 08/29/2023

Serial Number: 80-001263
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 20226C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 28021080A2 Exp: 12/05/2023
0.000	0.048	0.078	0.197	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.197	0.079

Number of Simulators Used: 5


Remarks:
NO COMMENTS

Lot # *
~~20206E~~
202206C
Missed the
"0" when
entering

-M. Wimmer

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MATTHEW S WIMMER
Signature and Printed Name

08/29/2023
Date