

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 13:14

Date of Inspection: 11/29/2023

Serial Number: 80-001311  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:WO32302427 Exp: 09/11/2024
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.081	0.199	0.080
0.000	0.051	0.081	0.200	0.079

Number of Simulators Used: 5

Remarks:

*In Compliance*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

W. P. May #F383  
Signature and Printed Name

WILLIAM P MAY

11/29/2023  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO

Time of Inspection: 20:04

Date of Inspection: 10/30/2023

Serial Number: 80-001311

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:WO32302427 Exp: 09/11/2024
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*William P May* #F583

WILLIAM P MAY

Signature and Printed Name

10/30/2023

Date



# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 10:06

Date of Inspection: 09/29/2023

Serial Number: 80-001311  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:WO32302427 Exp: 09/11/2024
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.051	0.080	0.199	0.078

Number of Simulators Used: 5

Remarks:

*In Compliance*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*William P May #F383*

WILLIAM P MAY

Signature and Printed Name

09/29/2023

Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 13:06

Date of Inspection: 08/31/2023

Serial Number: 80-001311  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:W032302427 Exp: 09/11/2024
0.000	0.050	0.078	0.199	0.077
0.000	0.051	0.079	0.199	0.077
0.000	0.051	0.080	0.199	0.077

Number of Simulators Used: 5

Remarks:

Back in Service From FDLE Annual  
In Compliance

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*W. P. May* #F383

WILLIAM P MAY  
Signature and Printed Name

08/31/2023  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 08:51

Date of Inspection: 06/30/2023

Serial Number: 80-001311  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.042 / 0.044			
0.000	0.043 / 0.044			
0.000	0.043 / 0.044			

Number of Simulators Used: 5

### Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:CONTROL TOLERANCE NOT BETWE  
EN .045 .055

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WILLIAM P MAY  
Signature and Printed Name

06/30/2023  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 09:39

Date of Inspection: 06/30/2023

Serial Number: 80-001311  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:WO32302427 Exp: 09/11/2024
0.000	0.048	0.078	0.197	0.074 / 0.076
0.000	0.048	0.079	0.197	0.074 / 0.076
0.000	0.048	0.079	0.197	0.074 / 0.076

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

Repeat on .05

Repeat on .08 Dry Gas

In Compliance  
Taken out of Service  
Sent to FDLE for Inspection  
Due to Control Tolerance Repeats

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

P/s William P May #F383

WILLIAM P MAY

Signature and Printed Name

06/30/2023  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 11:57

Date of Inspection: 05/31/2023

Serial Number: 80-001311  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:WO32302427 Exp: 09/11/2024
0.000	0.049	0.079	0.197	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.050	0.079	0.199	0.081

Number of Simulators Used: 5

Remarks:

*In Compliance*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WILLIAM P MAY

Signature and Printed Name

05/31/2023  
Date



# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 11:20

Date of Inspection: 04/30/2023

Serial Number: 80-001311  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:WO32302427 Exp: 09/11/2024
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.079	0.196	0.080
0.000	0.051	0.080	0.193	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

W. P. May #583

WILLIAM P MAY

Signature and Printed Name

04/30/2023  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 10:49

Date of Inspection: 03/31/2023

Serial Number: 80-001311  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:W032302427 Exp: 09/11/2024
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:

*In Compliance  
Back in Service from FDLE Annual*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*William P May #383*

WILLIAM P MAY

Signature and Printed Name

03/31/2023  
Date



# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO  
Time of Inspection: 12:07

Date of Inspection: 01/31/2023

Serial Number: 80-001311  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:WO32302427 Exp: 09/11/2024
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080

Number of Simulators Used: 5

Remarks:

*In Compliance*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*W. P. May #383*

WILLIAM P MAY

Signature and Printed Name

01/31/2023  
Date