

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO  
Time of Inspection: 09:36

Date of Inspection: 01/28/2023

Serial Number: 80-001344  
Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      |     | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 202112A<br>Exp: 12/07/2023 | 0.08g/210L Test (g/210L)<br>Lot#: 202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:<br>Exp: | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:<br>Exp: |
|----------------------------|--------------------------------------------------------------|--------------------------------------------------------------|-------------------------------------------|--------------------------------------------------------|
| 0.000                      | 0.051                                                        | 0.000 / 0.000                                                |                                           |                                                        |
| 0.000                      | 0.051                                                        | 0.000 / 0.000                                                |                                           |                                                        |
| 0.000                      | 0.052                                                        | 0.000 / 0.000                                                |                                           |                                                        |

Number of Simulators Used: 5

**Remarks:**

RECONNECT SIM, Control Outside Tolerance. Non-compliance: NO READING 08.

*INSTRUMENT NOT READING ANY VALUES DURING .08 TEST. INSPECTION ABORTED TO CHECK SIMULATOR & SOLUTION. SIMULATOR VAPOR LOCKED. PERFORMED AS EXPECTED ON SUBSEQUENT INSPECTIONS.*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*DEP. M. FELTOVIC*

DEP M FELTOVIC

Signature and Printed Name

01/28/2023  
Date