

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 09:52

Date of Inspection: 01/28/2023

Serial Number: 80-001346
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.049	0.000 / 0.000		
0.000	0.049	0.000		
0.000	0.049	0.000		

Number of Simulators Used: 5

Remarks:

08:CHANGE SOLUTION, Sequence Aborted. Non-compliance: BAD SIM.

INSTRUMENT NOT READING ANY VALUES DURING ,08 TEST. ABORTED INSPECTION TO CHECK SIMULATOR & SOLUTION. SIMULATOR VAPOR LOCKETS. PERFORMED AS EXPECTED ON SUBSEQUENT INSPECTIONS.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DEP. M. FELTOVIC 1965

DEP M FELTOVIC

Signature and Printed Name

01/28/2023
Date