

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS MAYPORT

Time of Inspection: 13:54

Date of Inspection: 06/15/2023

Serial Number: 80-001368

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2023 <i>6-14-2024</i> <i>RL</i>	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A4 Exp: 02/05/2025
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.078	0.200	0.080
0.000	0.049	0.078	0.198	0.080

Number of Simulators Used: 5

Remarks:

AMENDED INSPECTION INCORRECT YEAR WAS ENTERED ON THE 0.08 REFERENCE SOLUTION. REFERENCE SOLUTION BAR DATE 6-14-2024 NOT 6-14-2023

RL 6/16/24

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RL 6/16/24

Signature and Printed Name

ROBERT D LIGHTSEY

06/15/2023

Date