

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD

Serial Number: 80-002462

Time of Inspection: 13:52

Date of Inspection: 09/21/2023

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402370890 Exp: 05/09/2025
0.000	0.049	0.080	0.198	0.077
0.000	0.049	0.080	0.198	0.078
0.000	0.049	0.080	0.199	0.078

Number of Simulators Used: 5

Remarks:

RETIGHTENED SEALS AND LUBRICATED GASKETS. DEPARTMENT INSPECTOR NOTIFIED.
COMPLETE RETEST OF INSTRUMENT OK. 0.08g/210L TEST OK.

[Signature]
#1614

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]
#1614

CESAR CORREA

Signature and Printed Name

09/21/2023
Date