

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD

Time of Inspection: 10:44

Date of Inspection: 12/14/2023

Serial Number: 80-002464

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402370890 Exp: 05/09/2025
0.000	0.048	0.078	0.000 / 0.198	0.078
0.000	0.048	0.079	0.000 / 0.198	0.078
0.000	0.048	0.078	0.000 / 0.198	0.078

Number of Simulators Used: 5

Remarks:

SPOKE TO FDLE TAYLOR, RINSED SIMULATOR, PUT NEW SOLUTION, turned on Simulator, waited 30 minutes, and retested.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jose Montero Jr #1763

JOSE MONTERO

Signature and Printed Name

12/14/2023
Date