

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TALLAHASSEE PD

Time of Inspection: 14:41

Date of Inspection: 03/24/2023

Serial Number: 80-002989

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:33922080A2 Exp: 01/05/2025
0.000	0.044 / 0.045	0.075	0.196	0.080
0.000	0.045 / 0.046	0.075	0.196	0.081
0.000	0.045 / 0.045	0.075	0.196	0.081

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance.

- *fingerprints lid 3/24*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JJH TPD 321

JAMES J HARRIS

Signature and Printed Name

03/24/2023

Date