

Florida Department of Law Enforcement

Alcohol Testing Program

#1

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT WALTON BEACH PD
Time of Inspection: 00:30

Date of Inspection: 02/28/2023

Serial Number: 80-004770
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.000 / 0.044			
0.000	0.000 / 0.044			
0.000	0.000 / 0.044			

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: . Possible Bad Simulator
Changed Simulator for 0.05 solution and re-ran inspection.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cpl. Ben Crawford 2637

BENJAMIN D CRAWFORD

Signature and Printed Name

02/28/2023
Date

Florida Department of Law Enforcement

Alcohol Testing Program

#2

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT WALTON BEACH PD

Time of Inspection: 01:09

Date of Inspection: 02/28/2023

Serial Number: 80-004770

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211143 Exp: 08/05/2023
0.000	0.046	0.077	0.198	0.080
0.000	0.046	0.077	0.199	0.080
0.000	0.046	0.077	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cpl. Ben Crawford 2637

BENJAMIN D CRAWFORD

Signature and Printed Name

02/28/2023

Date