

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 11:55

Date of Inspection: 11/08/2023

Serial Number: 80-006475
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:402477283 Exp: 06/24/2025	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.048	0.000 / 0.000		
0.000	0.048	0.000 / 0.000		
0.000	0.048	0.000 / 0.000		

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. INSTRUMENT NOT READING GAS. . J GUZMAN IN TRAINING. Non-compliance: INSTRUMENT NOT READING DRY GAS..

Attached Dry Gas Standard Instead of .08g/210L Alcohol Reference Solution.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


ANAYA S FRAZIER
Signature and Printed Name

11/08/2023
Date

QND AMENDMENT

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 11:19

Date of Inspection: 10/04/2023

Serial Number: 80-006475
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202203J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 402477283 Exp: 06/24/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 402477283 Exp: 10/04/2023
0.000	0.048	0.077	0.000 / 0.198	0.079
0.000	0.048	0.077	0.000 / 0.198	0.078
0.000	0.048	0.077	0.000 / 0.199	0.079

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. ATTACHED WRONG SIM. J GUZMAN

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ANAYA S FRAZIER
Signature and Printed Name

10/04/2023
Date

AMENDMENT

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 11:19

Date of Inspection: 10/04/2023

Serial Number: 80-006475
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202203J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 402477283 Exp: 06/24/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 402477283 Exp: 10/04/2023
0.000	0.048	0.077	0.000 / 0.198	0.079
0.000	0.048	0.077	0.000 / 0.198	0.078
0.000	0.048	0.077	0.000 / 0.199	0.079

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. ATTACHED WRONG SIM. J GUZMAN

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ANAYA S FRAZIER

Signature and Printed Name

10/04/2023
Date



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Broward CSO

Instrument Serial Number: 80-006475

AGENCY INSPECTION DISCREPANCY

Agency Inspector: Anaya Frazier

Inspection Date: 09/06/2023

☐ A test was repeated without reason for repeat or the corrective action given

☐ Alcohol Free/Mouth Alcohol ☐ Alcohol Free Test ☐ Interferent Detect

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ Agency Inspection not conducted for calendar month

☐ Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☒ Lot Number and/or Expiration Date for Standards entered incorrectly

☒ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☒ Other Electronic Data Review/Comments: The 0.05 ARS lot number is mistyped.

REQUESTED CORRECTIVE ACTION

☒ Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector

☐ Provide written explanation regarding referenced item(s) to Department Inspector

☐ Remove instrument from evidentiary use until otherwise directed by the Department

☐ Other:

Taylor Gutschow

Digitally signed by Taylor

Gutschow

Date: 2023.10.31 14:33:23 -04'00'

10/31/2023

Signature of Alcohol Testing Program Staff Member

Date

Amendment
Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 12:21

Date of Inspection: 09/06/2023

Serial Number: 80-006475
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 2022033 Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 402477283 Exp: 06/24/2025
0.000	0.048	0.077	0.198	0.073 / 0.079
0.000	0.047	0.077	0.198	0.000 / 0.079
0.000	0.048	0.077	0.198	0.000 / 0.078

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. GAS NOT ATTACHED SECURELY. J. GUZMAN IN TRAINING.

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


ANAYA S FRAZIER
Signature and Printed Name

09/06/2023
Date