

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO  
Time of Inspection: 16:56

Date of Inspection: 03/10/2023

Serial Number: 80-006478  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08622080A1 Exp: 06/05/2024
0.000	0.052	0.081	0.202	0.079
0.000	0.052	0.082	0.202	0.079
0.000	0.052	0.081	0.202	0.079

Number of Simulators Used: 5

Remarks: LYSOL HAD BEEN HEAVILY USED BY THE BAT TECHNICIANS IN THEIR BATHROOM. I CLOSED THE BATHROOM DOOR TO BLOCK THE FUMES. PRIOR TO BLOWING THE ALCOHOL FREE SAMPLE, I COULD STILL SMELL LYSOL FUMES WHILE I WAS NEAR THE CLOSED BATHROOM DOOR RETRIEVING A MOUTH PIECE FROM THE COUNTER. THIS IN TURN CAUSED A .007 READING DURING THE ALCOHOL FREE BREATH SAMPLE, RESULTING IN A NO FOR THE ALCOHOL FREE BREATH SAMPLE SUBJECT TEST.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #6467

JASON C KARLECKE

Signature and Printed Name

03/10/2023  
Date