

# AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Instrument Serial Number: 80-006631 **Agency:** Manatee CSO AGENCY INSPECTION DISCREPENCY Inspection Date: 02/06/2023 Agency Inspector: William Coleman A test was repeated without reason for repeat or the corrective action given Alcohol Free/Mouth Alcohol ■ Alcohol Free Test Interferent Detect □ 0.080 g/210L □ 0.200 g/210L ☐ 0.080 g/210L Dry Gas □ 0.050 g/210L ☐ Agency Inspection not conducted for calendar month ☐ 0.080 g/210L Dry Gas □ 0.080 g/210L □ 0.200 g/210L □ 0.050 g/210L Lot Number and/or Expiration Date for Standards entered incorrectly □ 0.050 g/210L □ 0.080 g/210L □ 0.200 g/210L □ 0.080 g/210L Dry Gas Other Electronic Data Review/Comments: REQUESTED CORRECTIVE ACTION ■ Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector ☐ Provide written explanation regarding referenced item(s) to Department Inspector ☐ Remove instrument from evidentiary use until otherwise directed by the Department Other: Digitally signed by Taylor Taylor Gutschow Gutschow 03/21/2023 Date: 2023.03.21 20:02:44 -04'00' Signature of Alcohol Testing Program Staff Member **Date** 

## Florida Department of Law Enforcement **Alcohol Testing Program**

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO Time of Inspection: 09:35

Date of Inspection: 02/06/2023

Serial Number: 80-006631

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)		e	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402588895 Exp: 10/03/2025
RFI	/	0.000	0.049	0.079	0.198	0.080
	/	0.000	0.050	0.079	0.198	0.079
	/	0.000	0.050	0.079	0.198	0.080

Number of Simulators Used: 5

Ŕ	e	ma	r	ŀk	5	•	

Located source; corrected and re-ran test. (WDC) 00: RFI Detect.

Alcohol not present on breath during mouth alcohol test. Conducted a repeat and alcohol was detected. No further issues. (NOC)

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WILLIAM D COLEMAN

Signature and Printed Name

02/06/2023 Date

## Florida Department of Law Enforcement Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO Time of Inspection: 09:35

Date of Inspection: 02/06/2023

Serial Number: 80-006631

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		110
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)		è	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402588895 Exp: 10/03/2025
RFI	/	0.000	0.049	0.079	0.198	0.080
	/	0.000	0.050	0.079	0.198	0.079
	/	0.000	0.050	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:

00: RFI Detect. Located source; corrected and re-ran test. (WDC)

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WILLIAM D COLEMAN

Signature and Printed Name

02/06/2023 Date