

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO

Serial Number: 80-006694

Time of Inspection: 17:46

Date of Inspection: 11/06/2023

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 402558895 Exp: 10/03/2025
0.000	0.018 / 0.049	0.078	0.198	0.080
0.000	0.019 / 0.049	0.078	0.199	0.080
0.000	0.022 / 0.049	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance.

.05 Reference Solution heater shut off causing control outside tolerance; a 2nd test was run after heater sat for 10 minutes. Department Inspector notified - WOC

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WILLIAM COLEMAN

Signature and Printed Name

11/06/2023

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 13:46

Date of Inspection: 06/12/2023

Serial Number: 80-006694
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/14/2024 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402419647 Exp: 04/14/2025
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080

Number of Simulators Used: 5

Remarks:

Time-Date changed.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WILLIAM COLEMAN
Signature and Printed Name

06/12/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 14:28

Date of Inspection: 02/06/2023

Serial Number: 80-006694
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402419647 Exp: 04/14/2025
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.079	0.198	0.081
0.000	0.049	0.079	0.199	0.081


Number of Simulators Used: 5

Remarks:

Alcohol Free Test; corrected deficiency and re-ran test. (WOC)

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

WILLIAM D COLEMAN

02/06/2023
Date