

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER
Time of Inspection: 11:54

Date of Inspection: 01/11/2024

Serial Number: 80-006759
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.000 / 0.000			
0.000	0.000 / 0.000			
0.000	0.000 / 0.000			

Number of Simulators Used: 5

Remarks:

05: Control Outside FAILED BOTH TESTS READ 0.000. Non-compliance: During the monthly inspection the 0.05g/210L test solution was used for Instrument Serial # 80-006621. That instrument was in compliance and passed inspection. The same solution was used for the instrument with serial number 80-006759. All seals and hoses were checked as well as connections before hand. Instrument 80-006759 registered all zeros 0.000 during the control test. During the repeat test the result was the same.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISMAEL J TORRES

Signature and Printed Name

01/11/2024
Date