## Florida Department of Law Enforcement Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY S.O. Time of Inspection: 10:29

Date of Inspection: 10/04/2023

Serial Number: 80-007108

Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     |    |
|   |     | No |
| Diagnostic Check (Pre-Inspection): OK       |     |    |
|   | Yes |    |
| Alcohol Free Subject Test: 0.000            |     |    |
|   | Yes |    |
| Mouth Alcohol Test: Slope Not Met           |     |    |
|   | Yes |    |
| Interferent Detect Test: Interferent Detect |     |    |
|   | Yes |    |
| Diagnostic Check (Post-Inspection): OK      |     |    |
|   | Yes |    |

| Alcohol Free<br>Test<br>(g/210L) | 0.05g/210L Test<br>(g/210L)<br>Lot#: <del>202203J</del><br>Exp: 03/28/2025 | 0.08g/210L Test<br>(g/210L)<br>Lot#:202206B<br>Exp: 06/14/2024 | 0.20g/210L Test<br>(g/210L)<br>Lot#:202206C<br>Exp: 06/15/2024 | 0.08 g/210L<br>Dry Gas Std Test<br>(g/210L)<br>Lot#:402477283<br>Exp: 06/24/2025 |
|----------------------------------|--|--|--|--|
| 0.000                            | 0.049  | 0.079  | 0.199  | 0.082  |
| 0.000                            | 0.049  | 0.079  | 0.199  | 0.082  |
| 0.000                            | 0.049  | 0.080  | 0.200  | 0.081  |

Number of Simulators Used: 5

Remarks:

JGUZMAN IN TRAINING

A 19023 0.059/210 L TOST LOT #: 2023031

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that t hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

Date

FDLE/ATP Form 40 - March 2004



### AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Instrument Serial Number: 80-007108 **Agency:** Broward CSO AGENCY INSPECTION DISCREPENCY Inspection Date: 07/05/2023 **Agency Inspector:** Anaya Frazier A test was repeated without reason for repeat or the corrective action given Alcohol Free Test Interferent Detect Alcohol Free/Mouth Alcohol ☐ 0.080 g/210L Dry Gas □ 0.080 g/210L □ 0.200 g/210L □ 0.050 q/210L Agency Inspection not conducted for calendar month Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection ☐ 0.080 g/210L Dry Gas □ 0.080 g/210L □ 0.200 g/210L ☐ 0.050 g/210L Lot Number and/or Expiration Date for Standards entered incorrectly ■ 0.080 g/210L ■ 0.200 g/210L ■ 0.080 g/210L Dry Gas ■ 0.050 g/210L ■ Other Electronic Data Review/Comments: The lot numbers of the 0.05, 0.08, and 0.20 ARS are mistyped. The expiration date of the DGS is mistyped. REQUESTED CORRECTIVE ACTION Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector Provide written explanation regarding referenced item(s) to Department Inspector ☐ Remove instrument from evidentiary use until otherwise directed by the Department ☐ Other: Digitally signed by Taylor

Signature of Alcohol Testing Program Staff Member

Page 1 of 1

Date

10/30/2023

Taylor Gutschow Gutschow

Date: 2023,10.30 13:04:09 -04'00'

# 1end Ment

### Florida Department of Law Enforcement **Alcohol Testing Program**

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY S.O. Time of Inspection: 09:33

Date of Inspection: 07/05/2023

Serial Number: 80-007108 Software: 8100.27

|   | YES | NO |
|---|-----|----|
| Check or Test                               |     | -  |
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free<br>Test<br>(g/210L) | 0.05g/210L Test<br>(g/210L)<br>Lot#:202-112A<br>Exp: 12/07/2023 | 0.08g/210L Test<br>(g/210L)<br>Lot#:202-206B<br>Exp: 06/14/2024 | 0.20g/210L Test<br>(g/210L)<br>Lot#:202-206C<br>Exp: 06/15/2024 | 0.08 g/210L<br>Dry Gas Std Test<br>(g/210L)<br>Lot#:402477283<br>Exp:_07/05/2023 |
|----------------------------------|---|---|---|--|
| 0.000                            | 0.049   | 0.078   | 0.199   | 0.082  |
| 0.000                            | 0.049   | 0.078   | 0.200   | 0.082  |
| 0.000                            | 0.049   | 0.079   | 0.200   | 0.082  |

Number of Simulators Used: 5

Remarks:

LOT #: 202112A

D.08 9 LOT#: 202206.B

202206.C

.08 Dry Clas EXPIRATION DATE: 06/24/2025

The above inst

) does not comply (

) with Chapter 11D-8, FAC.

ANAYA S FRAZIER

Law Enforcement Agency Inspector Permit and that I I certify that with the provisions of Chapter 11D-8, FAC. performed this

07/05/2023

FDLE/ATP Form 40 - March 2004