

AMENDMENT

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 11:24

Date of Inspection: 10/04/2023

Serial Number: 80-007433
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202203J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 402477283 Exp: 06/24/2025
0.000	0.051	0.081	0.202	0.079
0.000	0.051	0.081	0.203	0.079
0.000	0.051	0.081	0.203	0.079

Number of Simulators Used: 5


Remarks:

J. GUZMAN IN TRAINING.

04
11/5/2023 0.05g/210L Test
LOT # 202303J

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ANAYA S FRAZIER
Signature and Printed Name

10/04/2023
Date



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Broward CSO

Instrument Serial Number: 80-007433

AGENCY INSPECTION DISCREPANCY

Agency Inspector: Anaya Frazier

Inspection Date: 07/05/2023

- ☐ A test was repeated without reason for repeat or the corrective action given
- ☐ Alcohol Free/Mouth Alcohol ☐ Alcohol Free Test ☐ Interferent Detect
- ☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas
- ☐ Agency Inspection not conducted for calendar month
- ☐ Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection
- ☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas
- ☒ Lot Number and/or Expiration Date for Standards entered incorrectly
- ☒ 0.050 g/210L ☒ 0.080 g/210L ☐ 0.200 g/210L ☒ 0.080 g/210L Dry Gas
- ☐ Other Electronic Data Review/Comments: The lot numbers of the 0.05 and 0.08 ARS are mistyped. The expiration date of the 0.08 DGS is mistyped as well.

REQUESTED CORRECTIVE ACTION

- ☒ Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector
- ☐ Provide written explanation regarding referenced item(s) to Department Inspector
- ☐ Remove instrument from evidentiary use until otherwise directed by the Department
- ☐ Other:

Taylor Gutschow

Digitally signed by Taylor

Gutschow

Date: 2023.10.30 13:02:19 -04'00'

Signature of Alcohol Testing Program Staff Member

10/30/2023

Date

AMENDMENT

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 12:16

Date of Inspection: 07/05/2023

Serial Number: 80-007433
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: 202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 402477283 Exp: 07/05/2023
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.050	0.081	0.203	0.080

Number of Simulators Used: 5

Remarks:
J. GUZMAN IN TRAINING.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 ANAYA S FRAZIER
Signature and Printed Name

07/05/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Broward County SO

Serial Number: 80-007433

Time of Inspection:

Date of Inspection: March 2023

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks: Inspection was not conducted as the instrument was sent to FDLE for Department Inspection and is awaiting part for replacement.

The above instrument complies (____) does not comply (____) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

03/11/2023

Date