

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 13:37

Date of Inspection: 11/02/2021

Serial Number: 80-007456

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:04120080A1 Exp: 03/05/2022
0.000	0.048	0.078	0.196	0.081
0.000	0.048	0.077	0.196	0.081
0.000	0.048	0.078	0.196	0.081

Number of Simulators Used: 5

Remarks:
NONE

NOTE:
I noticed on a Records Search, that I did not sign this document on 11/2/21. I signed it on 1/25/23. *[Signature]*

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 1/25/23

FRED C HENING

Signature and Printed Name

11/02/2021
Date