

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 13:21

Date of Inspection: 01/03/2023

Serial Number: 80-007492

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: 202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) 202206C Lot#: 202106B Exp: 06/22/2023 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31420080A2 Exp: 01/05/2023
0.000	0.048	0.078	0.198	0.077
0.000	0.048	0.078	0.198	0.077
0.000	0.047	0.078	0.198	0.077

Number of Simulators Used: 5

Remarks:

INSTRUMENT WAS CHECKED FOR DAMAGES. NONE WERE OBSERVED

0.20g/210L SOLUTION LOT# WAS TYPED INCORRECT. 01/03/2023

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ADAM B DESANTIS

Signature and Printed Name

01/03/2023

Date