Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FISH & WILDLIFE CC Time of Inspection: 14:35

Date of Inspection: 08/29/2023

Serial Number: 80-007502 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Үез	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402590345 Exp: 11/04/2025
0.000	0.049	0.079	0.197	0.081
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.078	0.197	0.081

Number of Simulators Used: 5

Remarks:

A F / M A:NOT ENOUGH MOUTH ALCOHOL SOLUTION.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name OSCAR SAAVEDRA

08/29/2023 Date

FDLE/ATP Form 40 - March 2004

['] Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FISH & WILDLIFE CC Time of Inspection: 13:30

Date of Inspection: 08/29/2023

Serial Number: 80-007502 Software: 8100.27

.

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK	ж	No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:

A F / K A:DID NOT HAVE SOLUTION IN MOUTH. Non-compliance:NO MOUTH ALCOHOL DURING MOUTH ALCOHOL SAMPL Z.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name • ۰. OSCAR SAAVEDRA 08/29/2023 Date

FDLE/ATP Form 40 - March 2004



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Fish & Wildlife CC	Instrument Serial Number: 80-007502
AGENCY INSPECTION DISCREPENCY	
Agency Inspector: Guillermo Cartaya	Inspection Date: 01/23/2023
	Alcohol Free Test Interferent Detect
-	Dry Gas Standard was used for inspection
□ 0.050 g/210L □ 0.080 g/210L □]0.200 g/210L
□ Lot Number and/or Expiration Date for □ 0.050 g/210L □ 0.050 g/210L	Standards entered incorrectly]0.200 g/210L
Other Electronic Data Review/Comment	ts:
REQUESTED CORRECTIVE ACTION	
Record hand-written amendments on F amendments, mark report "AMENDED"	DLE/ATP Form 40, initial and date ', and forward copy to Department Inspector
Provide written explanation regarding r	referenced item(s) to Department Inspector

Remove instrument from evidentiary use until otherwise directed by the Department

Other:

Taylor Gutschow Gutschow

Date: 2023.03.20 14:48:33 -04'00'

03/20/2023

Signature of Alcohol Testing Program Staff Member

Date