

Gutschow, Taylor

From: Gutschow, Taylor
Sent: Thursday, March 2, 2023 8:56 AM
To: 'Socarras, Pedro'
Subject: RE: Intoxilyzer 80-000879
Attachments: Form 40.pdf

Good morning,

Here is a Form 40 for you to fill out for the missed inspection. Please include a remark similar to what you said below... that the inspection could not be completed because the instrument would not power on and you have discussed with me and you are sending it to repair. When you send it to repair, please indicate there are records that need to be uploaded.

When you return the form to me, I'll get it added to the website.

If you have any further questions, let me know.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

From: Socarras, Pedro <PedroSocarras@miamibeachfl.gov>
Sent: Wednesday, March 1, 2023 11:17 AM
To: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>; nicolepierluisi@miamisao.com; Michael C Fente <MichaelCFente@MiamiSAO.com>
Cc: Price, David <DavidPrice@miamibeachfl.gov>; Rodriguez, Joaquin <JoaquinRodriguez@miamibeachfl.gov>; Millan, Kevin <KevinMillan@miamibeachfl.gov>
Subject: Intoxilyzer 80-000879

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Good morning,

I am writing to inform you that Intoxilyzer 80-000879 was taken out of service on Thursday, February 23rd, 2023. The Intoxilyzer is being sent for service due to an issue with it not wanting to power on. As a result, we are unable to complete an agency inspection or run any subject tests. Due to this power issue, an agency inspection was not conducted for February 2023 and all the data was not uploaded to FDLE.

Representatives from FDLE and the Miami SAO have been included in this notification. I will advise once the Intoxilyzer is back in service and please advise if there are any questions regarding this matter.

Regards,
PS

MIAMI BEACH

Pedro A. Socarras, Detective

Miami Beach Police Department

DUI Coordinator/Traffic Homicide Unit

1100 Washington Avenue, Miami Beach, FL 33139

Tel: (305)673-7776 ext. 5501 / Fax: (786)394-5279

pedrosocarras@miamibeachfl.gov

Mission: *Address Crime and Community Concerns*

Vision: *A safe and welcoming environment for everyone*

Values: *Honorable, Professional, Resilient*

Our Daily Goals: *Use innovative approaches to address crime, maintain and enhance a professional and well trained workforce, enhance the public's perception of the Miami Beach Police Department*