

Gutschow, Taylor

From: Kern, Michael <mkern@ccsofl.net>
Sent: Monday, April 4, 2022 2:58 AM
To: Gutschow, Taylor
Subject: Instruments 1739 and 7083

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

The following instruments are being sent to FDLE Fort Myers for further testing.

1739- System failure contact CMI error

7083- Ram failure on diagnostic will not come out of standby mode.

Sgt. Kern
Charlotte County Sheriff's Office.