

Gutschow, Taylor

From: Gutschow, Taylor
Sent: Thursday, June 15, 2023 3:24 PM
To: 'Closius, Richard L.'
Subject: 80-005550 Update

Update on instrument 80-005550. CMI was able to clear the error code and resolve the issue, but they could not retrieve any records (the instrument indicated it did not have any records on it). The June AI that you provided me should be uploaded to the website in the next week or two, along with this email as correspondence. If you have any breath tests run on the instrument during the affected timeframe, be advised that the data for them is not saved in COBRA, so I can't reprint them for you.

I will be inspecting the instrument in the next few days to make sure it is uploading and retaining records normally.

If you have any questions, feel free to reach out.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

From: Closius, Richard L. <rlclosius@mdpd.com>
Sent: Friday, June 2, 2023 10:38 AM
To: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>
Subject: RE: 80-005550

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Thank you.

RC

From: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>
Sent: Friday, June 2, 2023 10:29 AM

To: Closius, Richard L. <rlclosius@mdpd.com>

Cc: Myrtil, Reginald J. <rjmyrtil@mdpd.com>

Subject: RE: 80-005550

Good morning,

As we discussed earlier on the phone, please send the instrument to me for evaluation. I will see if I can pull the records manually. CMI can resolve the Error 12 remotely, and I believe the same is also true for the Error 10. If so, I will perform a Department Inspection on it as well.

As soon as you complete/return the loan agreement forms I sent you, I will send you a couple of loaner instruments.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

4700 Terminal Drive, Suite 3

Fort Myers, FL 33907

Desk #239-335-7500

Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

From: Closius, Richard L. <rlclosius@mdpd.com>

Sent: Friday, June 2, 2023 9:45 AM

To: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>

Cc: Myrtil, Reginald J. <rjmyrtil@mdpd.com>

Subject: 80-005550

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Taylor,

Ben was helping me upload instruments this morning. He thought he successfully uploaded 80-005550, but he noted that did not appear to be any tests on it. After the upload, I saw the "upload AI in 5 days" prompt still being displayed. I then attempted to get to the level two screen to check to see if the inspection was still on the instrument, but when I logged in, I got an Error 12/Contact CMI prompt. I was able to get into the level two menu, but there was nothing to recall. Ben tried connecting with the instrument again, but when he did, he saw that there were no records. I powered the instrument off, and back on, and repeated the process to see if my inspection was still there. This time, at log in, I got an Error 10/Contact CMI prompt.

I cannot reach the officer who was assigned to the instrument right now, but I will follow through and see if he had any BTOs on this instrument. I'll send you those if they exist.

Lastly, please send me the paperwork for a pair of loaner 8000s.

Thanks,

R. Closius

Gutschow, Taylor

From: Gutschow, Taylor
Sent: Friday, June 2, 2023 10:29 AM
To: 'Closius, Richard L.'
Cc: Myrtil, Reginald J.
Subject: RE: 80-005550

Good morning,

As we discussed earlier on the phone, please send the instrument to me for evaluation. I will see if I can pull the records manually. CMI can resolve the Error 12 remotely, and I believe the same is also true for the Error 10. If so, I will perform a Department Inspection on it as well.

As soon as you complete/return the loan agreement forms I sent you, I will send you a couple of loaner instruments.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

From: Closius, Richard L. <rlclosius@mdpd.com>
Sent: Friday, June 2, 2023 9:45 AM
To: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>
Cc: Myrtil, Reginald J. <rjmyrtil@mdpd.com>
Subject: 80-005550

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Taylor,

Ben was helping me upload instruments this morning. He thought he successfully uploaded 80-005550, but he noted that did not appear to be any tests on it. After the upload, I saw the "upload AI in 5 days" prompt still being displayed. I then attempted to get to the level two screen to check to see if the inspection was still on the instrument, but when I logged in, I got an Error 12/Contact CMI prompt. I was able to get into the level two menu, but there was nothing to recall. Ben tried connecting with the instrument again, but when he did, he saw that there were no records. I powered

the instrument off, and back on, and repeated the process to see if my inspection was still there. This time, at log in, I got an Error 10/Contact CMI prompt.

I cannot reach the officer who was assigned to the instrument right now, but I will follow through and see if he had any BTOs on this instrument. I'll send you those if they exist.

Lastly, please send me the paperwork for a pair of loaner 8000s.

Thanks,

R. Closius