

Gutschow, Taylor

From: Karlecke, Jason C. <KarleckeJ@pbso.org>
Sent: Friday, March 10, 2023 8:35 PM
To: Gutschow, Taylor; Nicodemo, Philip
Cc: Johnson, Jason R.; Fox, Adam H; Moody, Maria R
Subject: Two March 2023 Agency Inspections on 80-006478
Attachments: March 2023 AI for 80-006478.pdf

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Greetings,

On 3/10/2023, during the first March 2023 Agency Inspection of CMI Inc. Intoxilyzer 8000 S/N: 80-006478 I had a reading of .007 on the Alcohol Free Subject Test.

Just prior to blowing into 80-006478 I was standing next to the sink/counter area, just outside the BAT Technician's bathroom. Lysol had been heavily used in the BAT Technician's bathroom by the previous shift. I closed the bathroom door prior to starting my agency inspections in order to block the Lysol fumes. Prior to blowing the alcohol free sample for 80-006478, I could still smell Lysol fumes while I was near the closed bathroom when retrieving a mouth piece from the sink/counter area. This in turn caused a .007 reading during the Alcohol Free Subject Test. This occurred because I breathed in Lysol fumes that have an ethanol alcohol concentration 58.00% contaminating my breath/alcohol free sample.

I recognized what had occurred because this has happened before. I immediately moved my testing supplies away from BAT Technician's bathroom, then notified FDLE/ATP Quality Assurance Manager Philip Nicodemo of what had occurred via phone, advised Mr. Nicodemo that I would complete the current Agency Inspection of 80-006478 and conduct a retest on it to ensure it was functioning properly, in proper calibration, and in compliance with Florida Administrative Code (FAC) 11D-8. Mr. Nicodemo concurred with my assessment of what had occurred and agreed with my plan of action.

Upon completion of my first Agency Inspection of 80-006478, I verified the Alcohol Free Subject Test read NO when it should say YES, and retested 80-006478. As I took care to stay away from the BAT Technician's bathroom which was permeated with Lysol, 80-006478 correctly read 0.000 on the Alcohol Free Subject Test and reflected a YES in the Agency Inspection report (see attached March 2023 Agency Inspection reports for 80-006478).

Upon completion of my second Agency Inspection of 80-006478 I contacted Mr. Nicodemo and told him what had occurred. Mr. Nicodemo agreed that 80-006478 was in compliance with FAC 11D-8 and could be put into service for evidentiary breath collection.

I also notified BAT Unit Manager Maria Moody of what had occurred via phone.

Jason

Corporal Jason C. Karlecke #6467
Palm Beach Sheriff's Office

Agency Inspector/Unit Training Officer/D.U.I. Unit
3228 Gun Club Road
West Palm Beach, Florida 33406
Cell. (561) 660-1862
Email: KarleckeJ@pbso.org