



# Flow Calibration

FDLE  
Intoximeter - Alcotest Analyzer  
Model: 8000 SN: 80-000206  
09/20/2023  
Software: 8100.27

## Flow Rate Calibration\*\*\*\*\*

1: Rate (Liters/min) = 7.5

SQRT(Diff) = 7.277

2: Rate (Liters/min) = 15

SQRT(Diff) = 12.000

3: Rate (Liters/min) = 30

SQRT(Diff) = 21.281

Typical Flow Rate Error = 0.0000 L/min

Independent Data Scale Factor = 0.956





Rounded Slope = 688

Rounded Intercept = -7151.0

Correlation = 0.99749



# Stability Checks

| 0.05g/210L   | 0.08g/210L   | 0.20g/210L   | DGS 0.08g/210L   |
|--|--|--|--|
| 0.047 to 0.053   | 0.077 to 0.083   | 0.194 to 0.206   | 0.077 to 0.083   |
| ✓  | ✗  | ✓  | ✓  |
| <p>FILE<br/>C:\Users\A\Documents\210L\210L_05g\210L_05g_01.gpr<br/>Model: 8100<br/>SN: 81-000206<br/>09/20/2023<br/>Software: 8100.27</p> <p>Test Time</p> <p>Air Blank 0.000 11:20<br/>Control Test 0.049 11:21<br/>Air Blank 0.000 11:21<br/>Control Test 0.049 11:22<br/>Air Blank 0.000 11:22<br/>Control Test 0.049 11:23<br/>Air Blank 0.000 11:23</p> <p>Control Test Stats<br/>Average 0.0490<br/>Std Dev 0.0000<br/>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p> | <p>FILE<br/>C:\Users\A\Documents\210L\210L_08g\210L_08g_01.gpr<br/>Model: 8100<br/>SN: 81-000206<br/>09/20/2023<br/>Software: 8100.27</p> <p>Test Time</p> <p>Air Blank 0.000 11:26<br/>Control Test 0.079 11:27<br/>Air Blank 0.000 11:27<br/>Control Test 0.078 11:28<br/>Air Blank 0.000 11:28<br/>Control Test 0.076 11:29<br/>Air Blank 0.000 11:30</p> <p>Control Test Stats<br/>Average 0.0777<br/>Std Dev 0.0015<br/>Rel Std Dev(%) 1.9668</p> <p>Operator's Signature </p> | <p>FILE<br/>C:\Users\A\Documents\210L\210L_20g\210L_20g_01.gpr<br/>Model: 8100<br/>SN: 81-000206<br/>09/20/2023<br/>Software: 8100.27</p> <p>Test Time</p> <p>Air Blank 0.000 11:32<br/>Control Test 0.204 11:33<br/>Air Blank 0.000 11:34<br/>Control Test 0.205 11:35<br/>Air Blank 0.000 11:35<br/>Control Test 0.204 11:36</p> <p>Control Test Stats<br/>Average 0.2043<br/>Std Dev 0.0006<br/>Rel Std Dev(%) 0.2826</p> <p>Operator's Signature </p> | <p>FILE<br/>C:\Users\A\Documents\210L\210L_08g\210L_08g_02.gpr<br/>Model: 8100<br/>SN: 81-000206<br/>09/20/2023<br/>Software: 8100.27</p> <p>Test Time</p> <p>Air Blank 0.000 11:37<br/>Control Test 0.078 11:37<br/>Air Blank 0.000 11:37<br/>Control Test 0.079 11:38<br/>Air Blank 0.000 11:38<br/>Control Test 0.079 11:39<br/>Air Blank 0.000 11:39</p> <p>Control Test Stats<br/>Average 0.0787<br/>Std Dev 0.0006<br/>Rel Std Dev(%) 0.7339</p> <p>Operator's Signature </p> |





# Post-Cal Stability Checks

| 0.05g/210L   | 0.08g/210L   | 0.20g/210L   | DGS 0.08g/210L   |
|--|--|--|--|
| 0.047 to 0.053   | 0.077 to 0.083   | 0.194 to 0.206   | 0.077 to 0.083   |
| ✓  | ✓  | ✓  | ✓  |
| <p>FILE<br/>Packaging - Alcohol Challenge<br/>Model: 8000<br/>09/21/2023<br/>Software: 8100.27</p> <p>Test</p> <p>Time</p> <p>9/21/0L</p> <p>11:29</p> <p>11:29</p> <p>11:30</p> <p>11:30</p> <p>11:31</p> <p>11:31</p> <p>11:32</p> <p>11:32</p> <p>Control Test Stats</p> <p>Average 0.0503</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1471</p> | <p>FILE<br/>Packaging - Alcohol Challenge<br/>Model: 8000<br/>09/21/2023<br/>Software: 8100.27</p> <p>Test</p> <p>Time</p> <p>9/21/0L</p> <p>11:38</p> <p>11:38</p> <p>11:39</p> <p>11:39</p> <p>11:40</p> <p>11:40</p> <p>11:41</p> <p>11:41</p> <p>Control Test Stats</p> <p>Average 0.0780</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> | <p>FILE<br/>Packaging - Alcohol Challenge<br/>Model: 8000<br/>09/21/2023<br/>Software: 8100.27</p> <p>Test</p> <p>Time</p> <p>9/21/0L</p> <p>11:45</p> <p>11:45</p> <p>11:46</p> <p>11:46</p> <p>11:47</p> <p>11:47</p> <p>11:48</p> <p>11:48</p> <p>Control Test Stats</p> <p>Average 0.1997</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2892</p> | <p>FILE<br/>Packaging - Alcohol Challenge<br/>Model: 8000<br/>09/21/2023<br/>Software: 8100.27</p> <p>Test</p> <p>Time</p> <p>9/21/0L</p> <p>11:23</p> <p>11:23</p> <p>11:24</p> <p>11:24</p> <p>11:25</p> <p>11:25</p> <p>11:25</p> <p>11:25</p> <p>Control Test Stats</p> <p>Average 0.0793</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7277</p> |
| Operator's Signature   | Operator's Signature   | Operator's Signature   | Operator's Signature   |

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 14:28

Date of Inspection: 09/21/2023

Serial Number: 80-000206

Software: 8100.27

| Check or Test                                  | YES | NO | Check or Test                             | YES | NO |
|--|-----|----|---|-----|----|
| Diagnostic Check<br>(Pre-Inspection): OK       | Yes |    | Date and/or Time Adjusted                 |     | No |
| Minimum Sample Volume<br>Check: OK             | Yes |    | Barometric Pressure Sensor<br>Check: OK   | Yes |    |
| Alcohol Free Subject<br>Test: 0.000            | Yes |    | Mouth Alcohol Test:<br>Slope Not Met      | Yes |    |
| Interferent Detect Test:<br>Interferent Detect | Yes |    | Diagnostic Check<br>(Post-Inspection): OK | Yes |    |

| Alcohol Free<br>Test<br>(g/210L) | 0.05g/210L Test<br>(g/210L)<br>Lot#:202201C<br>Exp: 01/11/2024 | 0.08g/210L Test<br>(g/210L)<br>Lot#:202201D<br>Exp: 01/18/2024 | 0.20g/210L Test<br>(g/210L)<br>Lot#:202201E<br>Exp: 01/18/2024 | 0.08 g/210L<br>Dry Gas Std Test*<br>(g/210L)<br>Lot#:AG223802<br>Exp: 08/26/2024 |
|----------------------------------|--|--|--|--|
| 0.000                            | 0.050  | 0.078  | 0.200  | 0.079  |
| 0.000                            | 0.049  | 0.078  | 0.200  | 0.080  |
| 0.000                            | 0.049  | 0.078  | 0.200  | 0.080  |
| 0.000                            | 0.049  | 0.078  | 0.201  | 0.079  |
| 0.000                            | 0.049  | 0.078  | 0.199  | 0.079  |
| 0.000                            | 0.049  | 0.078  | 0.199  | 0.079  |
| 0.000                            | 0.049  | 0.077  | 0.199  | 0.079  |
| 0.000                            | 0.049  | 0.077  | 0.199  | 0.079  |
| 0.000                            | 0.049  | 0.078  | 0.200  | 0.080  |
| 0.000                            | 0.048  | 0.077  | 0.199  | 0.079  |

|                     |        |        |        |        |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0004 | 0.0006 | 0.0004 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/21/2023

Date





# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000206, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number: 80-000206  
Owning Agency: FDLE  
Calibration Date: 09/21/2023  
Calibration Time: 14:28

UNCERTAINTY\*  $\pm$

|                               |       |
|-------------------------------|-------|
| 0.050 g/210 L                 | 0.004 |
| 0.080 g/210 L                 | 0.004 |
| 0.200 g/210 L                 | 0.007 |
| 0.080 g/210 L Dry Gas Control | 0.005 |

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.  
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/21/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality