







INSTRUMENT PROCESSING SHEET

Agency Florida Department of Law EnforcementS/N 80-000227Florida Department of
Law EnforcementDate In 03/14/2023DI Completion Date 04/28/2023☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>04/06/2023</u>	Flow Calibration	By	Date																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>226</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.187</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
					Maintenance By _____																														
					<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____ _____																														
Calibration Adjustment By TDG _____					Department Inspection By TDG _____																														
Barometric Pressure Gauge <u>1015</u> ID # <u>28199</u>					Barometric Pressure ID# <u>26932</u>																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP4864</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>21410</td><td>09/30/2023</td></tr><tr><td>0.100</td><td>MP5099</td><td>22310</td><td>08/11/2024</td></tr><tr><td>0.200</td><td>MP5100</td><td>22050</td><td>02/07/2024</td></tr><tr><td>0.300</td><td>MP5101</td><td>22220</td><td>06/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904</td><td>06/08/2023</td></tr></tbody></table>					Simulator	Serial #	Lot #	Expiration	0.000	MP4864	N/A	N/A	0.040	MP5098	21410	09/30/2023	0.100	MP5099	22310	08/11/2024	0.200	MP5100	22050	02/07/2024	0.300	MP5101	22220	06/15/2024	0.080 DGS	N/A	AG115904	06/08/2023	Gauge <u>1012</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>		
Simulator	Serial #	Lot #	Expiration																																
0.000	MP4864	N/A	N/A																																
0.040	MP5098	21410	09/30/2023																																
0.100	MP5099	22310	08/11/2024																																
0.200	MP5100	22050	02/07/2024																																
0.300	MP5101	22220	06/15/2024																																
0.080 DGS	N/A	AG115904	06/08/2023																																
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096																
Simulator	Serial Number																																		
0.000	MP5092																																		
Interferent	MP5093																																		
0.050	MP5094																																		
0.080	MP5095																																		
0.200	MP5096																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C</td><td>01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D</td><td>01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E</td><td>01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802</td><td>08/26/2024</td></tr></tbody></table>					Simulator	Serial #	Lot #	Expiration	0.050	MP5094	202201C	01/11/2024	0.080	MP5095	202201D	01/18/2024	0.200	MP5096	202201E	01/18/2024	0.080 DGS	N/A	AG223802	08/26/2024	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____										
Simulator	Serial #	Lot #	Expiration																																
0.050	MP5094	202201C	01/11/2024																																
0.080	MP5095	202201D	01/18/2024																																
0.200	MP5096	202201E	01/18/2024																																
0.080 DGS	N/A	AG223802	08/26/2024																																
Notes/Suggested Service: <u>Instrument is an FDLE loaner and will be returned to evidentiary use when it is sent to a requesting agency. (TDG)</u> _____ _____ _____ _____ _____					<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																														
Israel Soto <small>Digitally signed by Israel Soto Date: 2023.05.01 14:18:51 +0000</small>					Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.05.03 11:20:32 -0400</small>																														
Tech Review / Date _____					Admin Review / Date _____																														

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000227	FDLE	04/06/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✗	✓	✗
<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 04/06/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:50</p> <p>Control Test 0.047 13:50</p> <p>Air Blank 0.000 13:51</p> <p>Control Test 0.048 13:51</p> <p>Air Blank 0.000 13:52</p> <p>Control Test 0.047 13:53</p> <p>Air Blank 0.000 13:53</p> <p>Control Test Stats</p> <p>Average 0.0473</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.2198</p> <p>Operator's Signature </p>	<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 04/06/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:58</p> <p>Control Test 0.076 13:58</p> <p>Air Blank 0.000 13:59</p> <p>Control Test 0.077 13:59</p> <p>Air Blank 0.000 14:00</p> <p>Control Test 0.077 14:01</p> <p>Air Blank 0.000 14:01</p> <p>Control Test Stats</p> <p>Average 0.0767</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7531</p> <p>Operator's Signature </p>	<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 04/06/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:05</p> <p>Control Test 0.198 14:05</p> <p>Air Blank 0.000 14:06</p> <p>Control Test 0.197 14:07</p> <p>Air Blank 0.000 14:07</p> <p>Control Test 0.198 14:08</p> <p>Air Blank 0.000 14:08</p> <p>Control Test Stats</p> <p>Average 0.1977</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2921</p> <p>Operator's Signature </p>	<p>DGS</p> <p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 04/06/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:42</p> <p>Control Test 0.076 13:42</p> <p>Air Blank 0.000 13:43</p> <p>Control Test 0.076 13:43</p> <p>Air Blank 0.000 13:44</p> <p>Control Test 0.076 13:44</p> <p>Air Blank 0.000 13:44</p> <p>Control Test Stats</p> <p>Average 0.0760</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>

Comments: Will perform an optical cal adjust. MG 04/06/2023

SCHEIDT

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000227	FDLE	04/27/2023	TDG ML

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 04/27/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:57</p> <p>Control Test 0.050 11:57</p> <p>Air Blank 0.000 11:58</p> <p>Control Test 0.049 11:59</p> <p>Air Blank 0.000 11:59</p> <p>Control Test 0.050 12:00</p> <p>Air Blank 0.000 12:00</p> <p>Control Test Stats</p> <p>Average 0.0497</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1625</p>	<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 04/27/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:04</p> <p>Control Test 0.079 12:05</p> <p>Air Blank 0.000 12:05</p> <p>Control Test 0.078 12:06</p> <p>Air Blank 0.000 12:06</p> <p>Control Test 0.079 12:07</p> <p>Air Blank 0.000 12:08</p> <p>Control Test Stats</p> <p>Average 0.0787</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7339</p>	<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 04/27/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:13</p> <p>Control Test 0.200 12:13</p> <p>Air Blank 0.000 12:14</p> <p>Control Test 0.200 12:14</p> <p>Air Blank 0.000 12:15</p> <p>Control Test 0.200 12:15</p> <p>Air Blank 0.000 12:16</p> <p>Control Test Stats</p> <p>Average 0.2000</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 04/27/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:51</p> <p>Control Test 0.079 11:51</p> <p>Air Blank 0.000 11:52</p> <p>Control Test 0.079 11:52</p> <p>Air Blank 0.000 11:52</p> <p>Control Test 0.079 11:52</p> <p>Air Blank 0.000 11:53</p> <p>Control Test Stats</p> <p>Average 0.0790</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>
<p>ML Operator's Signature</p>	<p>ML Operator's Signature</p>	<p>ML Operator's Signature</p>	<p>ML Operator's Signature</p>

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 12:20

Date of Inspection: 04/28/2023

Serial Number: 80-000227

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079


Standard Deviations	0.0005	0.0004	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/28/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000227, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000227</u>	UNCERTAINTY* \pm
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/28/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:20</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor D Gutschow
TAYLOR D GUTSCHOW,
Date 04/28/2023

Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality