



INSTRUMENT PROCESSING SHEET

Agency FDLES/N 80-000228Florida Department of
Law EnforcementDate In 1/17/2023DI Completion Date 1/18/2023☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE
*see Notes

Intake	By PN	Quality Checks	By PN	Date <u>1/18/2023</u>	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Return from Agency.</u> <u>(i.e., Evidential loaner)</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>183</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.164</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5084</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10/25/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C 01/11/2024	0.080	MP5089	202201D 01/18/2024	0.200	MP5084	202201E 01/18/2024	0.080 DGS	N/A	AG229803 10/25/2024																		
Simulator	Serial #	Lot #/Exp																																	
0.050	MP5088	202201C 01/11/2024																																	
0.080	MP5089	202201D 01/18/2024																																	
0.200	MP5084	202201E 01/18/2024																																	
0.080 DGS	N/A	AG229803 10/25/2024																																	
Calibration Adjustment		By _____		Department Inspection			By <u>PN</u>																												
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28421</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					Gauge <u>1016</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5084</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5084																
Simulator	Serial Number																																		
0.000	MP5086																																		
Interferent	MP5087																																		
0.050	MP5088																																		
0.080	MP5089																																		
0.200	MP5084																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Attachment #1</u>											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: <u>Inadvertently performed wet control test prior to 0.05 g/210L stability check. Performed 0.05 g/210L and internal printer did not work (see 'Other' attachment). All values were nominal on initial 0.05 test.</u> <u>*Instrument remained at FDLE subsequent to inspection completed on 1/18. PN</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>2023.01.</u>																															
				Israel Soto Digitally signed by Israel Soto Date: 2023.01.20 15:45:20 +05'00' 25 Tech Review / Date <u>11:49:20</u> Admin. Review Date <u>05/00</u>																															

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 13:56

Date of Inspection: 01/18/2023

Serial Number: 80-000228

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.078	0.203	0.077
0.000	0.049	0.078	0.203	0.077
0.000	0.048	0.078	0.203	0.076
0.000	0.048	0.078	0.203	0.077
0.000	0.048	0.078	0.203	0.076
0.000	0.049	0.078	0.203	0.077
0.000	0.048	0.079	0.203	0.077
0.000	0.048	0.079	0.203	0.077
0.000	0.048	0.079	0.203	0.077
0.000	0.048	0.079	0.204	0.076


Standard Deviations	0.0004	0.0005	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


PHIL NICODEMO

Signature and Printed Name

01/18/2023
 Date

FDLE

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000228

01/18/2023

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:06
Control Test	0.050	11:06
Air Blank	0.000	11:07
Control Test	0.049	11:08
Air Blank	0.000	11:08
Control Test	0.049	11:09
Air Blank	0.000	11:10
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	



Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000228
01/18/2023
Software: 8100.27

Test	g/210L	Time
-----	-----	-----
Air Blank	0.000	11:11
Control Test	0.078	11:11
Air Blank	0.000	11:12
Control Test	0.078	11:13
Air Blank	0.000	11:13
Control Test	0.078	11:14
Air Blank	0.000	11:15
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000228
01/18/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:16
Control Test	0.204	11:16
Air Blank	0.000	11:17
Control Test	0.202	11:18
Air Blank	0.000	11:18
Control Test	0.202	11:19
Air Blank	0.000	11:20
Control Test Stats		
Average	0.2027	
Std Dev	0.0012	
Rel Std Dev(%)	0.5698	


Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000228
01/18/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:22
Control Test	0.077	11:22
Air Blank	0.000	11:23
Control Test	0.078	11:23
Air Blank	0.000	11:24
Control Test	0.078	11:24
Air Blank	0.000	11:24
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

D65



Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000228, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000228</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/18/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:56</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Phil Nicodem
Nicodemo
Date: 2023.01.18 15:42:29
-05'00'

Digitally signed by Phil

Nicodemo

Date: 2023.01.18 15:42:29
-05'00'

Date

PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

ATTACHMENT #1

WET CONTROL + STABILITY

