



Agency Escambia County Sheriff's Office

S/N 80-000744

Date In 06/09/2023 DI Completion Date 06-13-2023

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake		By <u>ALL</u>		Date <u>06-09-2023</u>																				
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>214</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks																						
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td rowspan="2">0.050</td><td rowspan="2">MP6291</td><td>202201C</td></tr><tr><td>01-11-2024</td></tr><tr><td rowspan="2">0.080</td><td rowspan="2">MP6292</td><td>202201D</td></tr><tr><td>01-18-2024</td></tr><tr><td rowspan="2">0.200</td><td rowspan="2">MP6293</td><td>202201E</td></tr><tr><td>01-18-2024</td></tr><tr><td rowspan="2">0.080 DGS</td><td rowspan="2">N/A</td><td>AG229803</td></tr><tr><td>10-25-2024</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C	01-11-2024	0.080	MP6292	202201D	01-18-2024	0.200	MP6293	202201E	01-18-2024	0.080 DGS	N/A	AG229803	10-25-2024		
Simulator	Serial #	Lot #/Exp																						
0.050	MP6291	202201C																						
		01-11-2024																						
0.080	MP6292	202201D																						
		01-18-2024																						
0.200	MP6293	202201E																						
		01-18-2024																						
0.080 DGS	N/A	AG229803																						
		10-25-2024																						
		<table border="1"><thead><tr><th>Flow Calibration</th><th>By <u>IS</u></th><th>Date</th></tr></thead><tbody><tr><td colspan="3">Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)</td></tr></tbody></table>		Flow Calibration	By <u>IS</u>	Date	Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																	
Flow Calibration	By <u>IS</u>	Date																						
Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																								
		<table border="1"><thead><tr><th>Maintenance</th><th>By <u>IS</u></th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Internal printer paper replaced 06-09-2023.</u></td></tr><tr><td colspan="2">_____</td></tr><tr><td colspan="2">_____</td></tr><tr><td colspan="2">_____</td></tr></tbody></table>		Maintenance	By <u>IS</u>	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Internal printer paper replaced 06-09-2023.</u>		_____		_____		_____												
Maintenance	By <u>IS</u>																							
☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Internal printer paper replaced 06-09-2023.</u>																								

Calibration Adjustment				By <u>IS</u>	
Barometric Pressure Gauge <u>1009</u> ID # <u>28662</u>				Department Inspection	
Simulator	Serial #	Lot #	Expiration	Barometric Pressure ID# <u>28421/28421</u>	
0.000	MP5091	N/A	N/A	Gauge <u>1008/1011</u> Instrument <u>1006/1009</u>	
0.040	MP5082	22460	12-28-2024	Mouth Alcohol Solution Lot # <u>2022-A</u>	
0.100	MP6296	22430	11-30-2024	Acetone Stock Solution Lot # <u>2022-B</u>	
0.200	MP5084	22400	10-12-2024		
0.300	MP5085	23070	03-06-2025		
0.080 DGS	N/A	06723080A5	04-05-2025		
☒ Post Calibration Adjustment Stability Checks					
Simulator	Serial #	Lot #	Expiration		
0.050	MP6291	202201C	01-11-2024		
0.080	MP6292	202201D	01-18-2024		
0.200	MP6293	202201E	01-18-2024		
0.080 DGS	N/A	AG229803	10-25-2024		
Notes/Suggested Service: Instrument had Interferent Detect on wet 0.080 g/210L test during Department Inspection. Optical Bench Calibration Adjustment was performed. Second Department Inspection performed, all results in nominal range. Mouth Alcohol/Acetone stock solution lots, and simulator serial #'s are the same for both Department Inspections. IS					

Attachments	
☒ Form 41x2 ☒ Stability Checks ☒ Calibration Certificate x2 ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____

☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC	
☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use	
☒ Conduct an Agency Inspection Before Evidentiary Use	

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2023.06.13 12:15:02 -0400	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.06.14 09:24:00 -0400
Tech Review / Date	Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 08:17

Date of Inspection: 06/13/2023

Serial Number: 80-000744
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.079	0.198	0.078
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.046	0.079	0.198	0.079
0.000	0.046	0.079	0.197	0.080
0.000	0.045	0.079	0.196	0.079
0.000	0.046	0.079	0.195	0.079
0.000	0.046	0.079	0.194	0.079
0.000	0.046	0.080	0.196	0.079

Standard Deviations	0.0011	0.0004	0.0014	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

*Wet 0.080g/210L
Lot#: 202303L
Exp: 03-29-2025*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

06/13/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 08:43

Date of Inspection: 06/12/2023

Serial Number: 80-000744
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.078 / 0.079	0.201	0.080
0.000	0.048	0.079 / 0.079	0.202	0.080
0.000	0.048	0.079 / 0.080	0.202	0.079
0.000	0.048	0.077 / 0.079	0.201	0.079
0.000	0.047	0.077 / 0.079	0.201	0.080
0.000	0.048	INT / 0.080	0.201	0.080
0.000	0.048	/ 0.080	0.201	0.079
0.000	0.048	/ 0.080	0.202	0.080
0.000	0.045	/ 0.080	0.201	0.080
0.000	0.045	/ 0.081	0.201	0.079
Standard Deviations	0.0012	/ 0.0006	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

08: Interferent Detect CHECKED SIM AND AREA

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

06/12/2023
Date

stability checks

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:26
Control Test	0.048	14:27
Air Blank	0.000	14:28
Control Test	0.048	14:28
Air Blank	0.000	14:29
Control Test	0.048	14:30
Air Blank	0.000	14:30
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:32
Control Test	0.078	14:32
Air Blank	0.000	14:33
Control Test	0.078	14:33
Air Blank	0.000	14:34
Control Test	0.078	14:35
Air Blank	0.000	14:35
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet


Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:37
Control Test	0.200	14:38
Air Blank	0.000	14:38
Control Test	0.199	14:39
Air Blank	0.000	14:40
Control Test	0.198	14:40
Air Blank	0.000	14:41
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	


Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:42
Control Test	0.079	14:42
Air Blank	0.000	14:43
Control Test	0.080	14:43
Air Blank	0.000	14:44
Control Test	0.080	14:44
Air Blank	0.000	14:44
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Dry


Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000744 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000744</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ESCAMBIA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/13/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:17</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2023.06.13 10:56:30 -04'00'

06/13/2023

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000744 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000744</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ESCAMBIA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/12/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:43</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2023.06.12 10:41:01 -04'00'

06/12/2023

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/12/2023 10:05:20

Optical Bench
Calibration Adjustment
✓

Auto Calibration

pg 1 of 2

	<<<<<	3um	>>>>>	<<<<<	9um	>>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.0470	(-0.0220)		0.1410	(-0.0100)	
Sample #2	0.0450	(0.0100)		0.1320	(0.0060)	
Sample #3	0.0590	(0.0450)		0.1450	(0.0270)	
Sample #4	0.0560	(0.0620)		0.1450	(0.0320)	
Avg % Abs	0.0533	(0.0390)		0.1407	(0.0217)	
STD DEV	0.0074	(0.0265)		0.0075	(0.0138)	
REL STD DEV	13.821	(67.985)		5.336	(63.674)	

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.7600	(-0.0160)		1.5620	(0.0020)	
Sample #2	0.7770	(-0.0200)		1.5690	(0.0000)	
Sample #3	0.7730	(0.0350)		1.5350	(0.0560)	
Sample #4	0.8020	(0.0060)		1.5670	(0.0310)	
Avg % Abs	0.7840	(0.0070)		1.5570	(0.0290)	
STD DEV	0.0157	(0.0275)		0.0191	(0.0281)	
REL STD DEV	2.005	(393.052)		1.225	(96.736)	

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	1.9020	(-0.0550)		3.6550	(-0.0290)	
Sample #2	1.8310	(-0.0030)		3.6130	(0.0260)	
Sample #3	1.8510	(0.0060)		3.6210	(0.0180)	
Sample #4	1.8150	(0.0160)		3.5720	(0.0440)	
Avg % Abs	1.8323	(0.0063)		3.6020	(0.0293)	
STD DEV	0.0180	(0.0095)		0.0263	(0.0133)	
REL STD DEV	0.984	(150.069)		0.730	(45.398)	

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	3.6060	(-0.0060)		6.9270	(-0.0120)	
Sample #2	3.5760	(0.0240)		6.8470	(0.0530)	
Sample #3	3.5530	(0.0330)		6.8170	(0.0750)	
Sample #4	3.5420	(0.0440)		6.8360	(0.0740)	
Avg % Abs	3.5570	(0.0337)		6.8333	(0.0673)	
STD DEV	0.0173	(0.0100)		0.0152	(0.0124)	
REL STD DEV	0.488	(29.752)		0.222	(18.450)	

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	5.2890	(-0.0250)		10.0280	(-0.0160)	
Sample #2	5.2450	(0.0610)		9.9520	(0.0960)	
Sample #3	5.2330	(0.0680)		9.9190	(0.1230)	
Sample #4	5.2400	(0.0430)		9.9270	(0.0900)	
Avg % Abs	5.2393	(0.0573)		9.9327	(0.1030)	
STD DEV	0.0060	(0.0129)		0.0172	(0.0176)	
REL STD DEV	0.115	(22.495)		0.173	(17.066)	

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/12/2023 10:05:20

Optical Bench
Calibration Adjustment
✓

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -154.81
First Order Coef 2639.06
Second Order Coef 22.47

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0003
0.040	0.040	-0.0005
0.100	0.100	0.0001
0.200	0.200	0.0002
0.300	0.300	-0.0001

<<<< 9um >>>>

-196.65
1335.01
12.44

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0002
0.040	0.040	-0.0002
0.100	0.100	-0.0002
0.200	0.200	0.0004
0.300	0.300	-0.0001

<<<< 3um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample

Sample #1	3476.00
Sample #2	3288.00
Sample #3	3452.00
Sample #4	3417.00
Avg	3385.6667
STD DEV	86.3732
REL STD DEV	2.551
H2O adjust (mg/l*10k)	424

<<<< 9um >>>>

3385.00
3339.00
3377.00
3389.00
3368.3333
26.1024
0.775
441

Barometric Pressure = 1009

*****CALIBRATION SUCCESSFUL*****

Post stability checks

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:48
Control Test	0.049	10:49
Air Blank	0.000	10:49
Control Test	0.048	10:50
Air Blank	0.000	10:50
Control Test	0.048	10:51
Air Blank	0.000	10:52
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:53
Control Test	0.079	10:53
Air Blank	0.000	10:54
Control Test	0.077	10:54
Air Blank	0.000	10:55
Control Test	0.078	10:56
Air Blank	0.000	10:56
Control Test Stats		
Average	0.0780	
Std Dev	0.0010	
Rel Std Dev(%)	1.2821	

wet

Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:58
Control Test	0.200	10:58
Air Blank	0.000	10:59
Control Test	0.197	10:59
Air Blank	0.000	11:00
Control Test	0.197	11:01
Air Blank	0.000	11:01
Control Test Stats		
Average	0.1980	
Std Dev	0.0017	
Rel Std Dev(%)	0.8748	

Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000744
06/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:02
Control Test	0.079	11:03
Air Blank	0.000	11:03
Control Test	0.079	11:04
Air Blank	0.000	11:04
Control Test	0.079	11:04
Air Blank	0.000	11:05
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry

Operator's Signature