



Agency Lafayette County Sheriff's Office

S/N 80-000776

Date In 03-27-2023 DI Completion Date 04-04-2023

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

03 IS 04-04-2023

Calibration Adjustment

By

Barometric Pressure Gauge _____ ID # _____

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: Tech Review: Corrected date on
Quality Checks. IS 04-04-2023

Department Inspection

By IS

Barometric Pressure ID# 30793

Gauge 1014	Instrument 1007
------------	-----------------

Mouth Alcohol Solution Lot # 2022-A

Acetone Stock Solution Lot # 2022-B

Attachments

- | | |
|---|---|
| ☒ Form 41
☒ Stability Checks
☒ Calibration Certificate
☐ Calibration Adjustment | ☐ Post-Stability Checks
☐ Flow Calibration
☐ Form 40
☐ Other |
|---|---|

- ☒ Instrument Complies with Chapter 11D-8, FAC
 - ☐ Instrument Does Not Comply with Chapter 11D-8, FAC

 - ☒ Return to/Place into Evidentiary Use
 - ☐ Remain Out of Evidentiary Use

 - ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2023.04.04 11:06:31
-04'00'

Phil Nicodemo

Digitally signed by Phil Nicodemo
Date: 2023.04.05 13:59:49 -04'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT – INTOXILYZER 8000

Agency: LAFAYETTE CO SO
Time of Inspection: 10:20

Date of Inspection: 04/04/2023

Serial Number: 80-000776
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.078	0.201	0.081
0.000	0.049	0.078	0.202	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.202	0.081
0.000	0.049	0.078	0.201	0.080
0.000	0.048	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.079	0.200	0.079


Standard Deviations	0.0004	0.0004	0.0010	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


ISRAEL SOTO
Signature and Printed Name

04/04/2023
Date

Stability checks

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000776
04/03/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:33
Control Test	0.049	10:34
Air Blank	0.000	10:35
Control Test	0.049	10:35
Air Blank	0.000	10:36
Control Test	0.049	10:36
Air Blank	0.000	10:37
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000776
04/03/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:38
Control Test	0.079	10:39
Air Blank	0.000	10:40
Control Test	0.077	10:40
Air Blank	0.000	10:41
Control Test	0.078	10:41
Air Blank	0.000	10:42
Control Test Stats		
Average	0.0780	
Std Dev	0.0010	
Rel Std Dev(%)	1.2821	

wet

Operator's Signature

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000776
04/03/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:46
Control Test	0.203	10:47
Air Blank	0.000	10:47
Control Test	0.201	10:48
Air Blank	0.000	10:49
Control Test	0.201	10:49
Air Blank	0.000	10:50
Control Test Stats		
Average	0.2017	
Std Dev	0.0012	
Rel Std Dev(%)	0.5726	

Operator's Signature

LAFAYETTE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000776
04/03/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:53
Control Test	0.078	10:53
Air Blank	0.000	10:54
Control Test	0.079	10:54
Air Blank	0.000	10:55
Control Test	0.079	10:55
Air Blank	0.000	10:56
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Dry

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000776 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000776</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LAFAYETTE CO SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/04/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:20</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2023.04.04 10:28:24 -04'00'

04/04/2023

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality