




INSTRUMENT PROCESSING SHEET

Agency Cape Coral Police DepartmentS/N 80-000859Florida Department of
Law EnforcementDate In 1/12/2023 DI Completion Date 1/17/2023☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DERR	Date 1/17/2023	Flow Calibration	By	Date																											
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>204</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP6287</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>00521080A2 02/05/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023																	
Simulator	Serial #	Lot #/Exp																																
0.050	MP6286	202201C 01/11/2024																																
0.080	MP6287	202201D 01/18/2024																																
0.200	MP6288	202201E 01/18/2024																																
0.080 DGS	N/A	00521080A2 02/05/2023																																
Calibration Adjustment		Department Inspection																																
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28663</u>		By DERR _____																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1019</u> Instrument <u>1017</u>				
Simulator	Serial #	Lot #	Expiration																															
0.000		N/A	N/A																															
0.040																																		
0.100																																		
0.200																																		
0.300																																		
0.080 DGS	N/A																																	
		Mouth Alcohol Solution Lot # <u>2021-D</u>																																
		Acetone Stock Solution Lot # <u>2021-C</u>																																
☐ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																			
Simulator	Serial Number																																	
0.000	MP6284																																	
Interferent	MP6285																																	
0.050	MP6286																																	
0.080	MP6287																																	
0.200	MP6288																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments												
Simulator	Serial #	Lot #	Expiration																															
0.050																																		
0.080																																		
0.200																																		
0.080 DGS	N/A																																	
		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment		☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																														
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																
		Israel Soto Digitally signed by Israel Soto Date: 2023.01.17 15:27:28 +0500		2023.01 16:56:1																														
		Tech Review / Date		Admin Review / Date																														



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000859, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000859</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>CAPE CORAL PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/17/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:35</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

01/17/2023

Date

David E Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CAPE CORAL PD
Time of Inspection: 12:35

Date of Inspection: 01/17/2023

Serial Number: 80-000859
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.048	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.078	0.199	0.080


Standard Deviations	0.0005	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA
Signature and Printed Name

01/17/2023
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000859	Cape Coral Police Department	01/17/2023	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 01/17/2023 Software: 8100.27 SN 80-000859 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:41</td></tr><tr><td>Control Test</td><td>0.049</td><td>09:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:43</td></tr><tr><td>Control Test</td><td>0.049</td><td>09:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:44</td></tr><tr><td>Control Test</td><td>0.049</td><td>09:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:45</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:41	Control Test	0.049	09:42	Air Blank	0.000	09:43	Control Test	0.049	09:43	Air Blank	0.000	09:44	Control Test	0.049	09:44	Air Blank	0.000	09:45	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 01/17/2023 Software: 8100.27 SN 80-000859 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:46</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:47</td></tr><tr><td>Control Test</td><td>0.078</td><td>09:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr><tr><td>Control Test</td><td>0.077</td><td>09:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2821</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:46	Control Test	0.079	09:47	Air Blank	0.000	09:47	Control Test	0.078	09:48	Air Blank	0.000	09:49	Control Test	0.077	09:49	Air Blank	0.000	09:50	Control Test Stats			Average	0.0780		Std Dev	0.0010		Rel Std Dev(%)	1.2821		CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 01/17/2023 Software: 8100.27 SN 80-000859 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:52</td></tr><tr><td>Control Test</td><td>0.199</td><td>09:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:53</td></tr><tr><td>Control Test</td><td>0.199</td><td>09:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:54</td></tr><tr><td>Control Test</td><td>0.198</td><td>09:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:55</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1987</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:52	Control Test	0.199	09:52	Air Blank	0.000	09:53	Control Test	0.199	09:53	Air Blank	0.000	09:54	Control Test	0.198	09:55	Air Blank	0.000	09:55	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 01/17/2023 Software: 8100.27 SN 80-000859 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:56</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:58</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:56	Control Test	0.080	09:57	Air Blank	0.000	09:57	Control Test	0.080	09:57	Air Blank	0.000	09:58	Control Test	0.080	09:58	Air Blank	0.000	09:59	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:41																																																																																																																																																	
Control Test	0.049	09:42																																																																																																																																																	
Air Blank	0.000	09:43																																																																																																																																																	
Control Test	0.049	09:43																																																																																																																																																	
Air Blank	0.000	09:44																																																																																																																																																	
Control Test	0.049	09:44																																																																																																																																																	
Air Blank	0.000	09:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:46																																																																																																																																																	
Control Test	0.079	09:47																																																																																																																																																	
Air Blank	0.000	09:47																																																																																																																																																	
Control Test	0.078	09:48																																																																																																																																																	
Air Blank	0.000	09:49																																																																																																																																																	
Control Test	0.077	09:49																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2821																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:52																																																																																																																																																	
Control Test	0.199	09:52																																																																																																																																																	
Air Blank	0.000	09:53																																																																																																																																																	
Control Test	0.199	09:53																																																																																																																																																	
Air Blank	0.000	09:54																																																																																																																																																	
Control Test	0.198	09:55																																																																																																																																																	
Air Blank	0.000	09:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2906																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:56																																																																																																																																																	
Control Test	0.080	09:57																																																																																																																																																	
Air Blank	0.000	09:57																																																																																																																																																	
Control Test	0.080	09:57																																																																																																																																																	
Air Blank	0.000	09:58																																																																																																																																																	
Control Test	0.080	09:58																																																																																																																																																	
Air Blank	0.000	09:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		